Determination of conjugated bile acids in human bile by isotachophoresis in a non-aqueous solvent using a.c. conductivity and UV detection

Citation for published version (APA):

Reijenga, J. C., Slaats, H. J. L. A., & Everaerts, F. M. (1983). Determination of conjugated bile acids in human bile by isotachophoresis in a non-aqueous solvent using a.c. conductivity and UV detection. *Journal of Chromatography*, 267(1), 85-89. [https://doi.org/10.1016/S0021-9673\(01\)90821-0](https://doi.org/10.1016/S0021-9673(01)90821-0)

DOI:

[10.1016/S0021-9673\(01\)90821-0](https://doi.org/10.1016/S0021-9673(01)90821-0)

Document status and date:

Published: 01/01/1983

Document Version:

Publisher's PDF, also known as Version of Record (includes final page, issue and volume numbers)

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.tue.nl/taverne

Take down policy

If you believe that this document breaches copyright please contact us at:

openaccess@tue.nl

providing details and we will investigate your claim.

CHROM. 15,923

DETERMINATION OF CONJUGATED BILE ACIDS IN HUMAN BILE BY ISOTACHOPHORESIS IN A NON-AQUEOUS SOLVENT USING a.c. CONDUCTIVITY AND UV DETECTION

J. C. REIJENGA*, H. J. L. A. SLAATS and F. M. EVERAERTS

Laboratory of Instrumental Analysis, Eindhoven University of Technology, P.O. Box 513, 5600 MB Eindhoven (The Netherlands)

(Received April 13th, 1983)

SUMMARY

A method for the determination of conjugated bile acids in human bile using isotachophoresis in 95 % methanol is described. The leading ion is 0.01 *M* chloride, the counter ion is hydroxylamine at its pK^* value and the terminating ion is N-2-hydroxyethylpiperazine-N'-2-ethanesulphonic acid (HEPES). The sample preparation consists of C_{18} -silica cartridge adsorption. Microlitre amounts of the methanol eluate are injected and analysed within 20 min in a 0.2 mm I.D. PTFE capillary. The sensitivity of the method is better than 50 ng of each of the conjugated bile acids using a.c. conductivity detection.

INTRODUCTION

Methods for the determination of bile acids have shown some improvements in the last decade. In a review in 1973¹, enzymatic methods were recommended for routine use, whereas a combination of thin-layer and gas-liquid chromatography was suggested if specific information on individual bile acids is needed. Although some improvements in enzymatic², thin-layer chromatographic³ and gas chromatographic⁴ methods have been reported recently, the developments in high-performance liquid chromatography (HPLC) are apparently more promising. In a more recent review⁵ on the separation of bile acids by HPLC, the retention behaviour of free and conjugated bile acids in both reversed-phase and normal-phase systems was reported. For the isolation of bile acids from biological material (bile, serum, faeces), sample pre-treatment is usually time consuming (from 10 min to more than 24 h), as a group separation of free and conjugated acids is generally included.

The detection of bile acids in HPLC is mostly carried out by differential refractive index measurements^{5, 8}. Some mobile phases allow UV detection at wavelengths below 210 nm^{8, 9}. Fluorescence detection with pre-column¹⁰ or post-column¹¹ derivatization has also been reported.

Another analytical technique that has developed significantly during the last decade is capillary isotachophoresis^{12–14}. The equipment uses a sensitive universal

detector compared with refractive index detection in HPLC, a.c. conductivity or potential gradient detection. Most applications are performed in aqueous operational systems, but the importance of using non-aqueous solvents was soon recognized^{12,15}. One reason is that the use of the technique can be extended to ionizable compounds, sparingly soluble in water, such as long-chain fatty acids¹⁵. Another reason is the use of specific solvent-solute interactions to optimize certain separations. As the solubility of bile acids under aqueous isotachophoretic conditions is insufficient, a methanolic operational system is applied for the analysis of free and glyco- and tauro-conjugated bile acids in biological samples.

EXPERIMENTAL

The use of non-aqueous (*e.g.*, methanolic) systems in isotachophoresis has so far been limited by the availability of insulating material that is both resistant to these solvents and suitable for the construction of a high-resolution a.c. conductivity detector, for which Perspex is usually used¹². The use of Teflon® has also been reported¹⁶. The introduction of an Araldite that meets both of these requirements was reported recently¹⁷. The construction of the equipment used was similar to that described earlier¹². The separation compartment consisted of a PTFE capillary of I.D. 0.2 mm and length *ca.* 25 cm. The a.c. conductivity cell had a volume of 3 nl, corresponding to a theoretical minimum detectable amount of *ca.* $30 \cdot 10^{-12}$ equivalents. UV detection was carried out at 206 nm.

The solvent used in the isotachophoretic experiments was 95 % (v/v) methanol in water. A higher methanol content is not necessary. A lower methanol content decreases the solubility of the bile acids to below the isotachophoretic steady-state concentration. The use of non-aqueous solvents in isotachophoresis is hampered by the lack of data on *pK*, mobility and solubility values of ions normally used in operational systems. As the leading ion chloride was used, the mobility of which is considerably lower than in water. For this particular application, however, this was irrelevant. As the counter ion we preferred buffers normally used in aqueous isotachophoresis, mainly because of their availability and relative purity. However, histidine, used preferably for neutral anionic systems, could not be used because of its limited solubility at methanol concentrations above 60 %. We used either hydrox-

TABLE I

OPERATIONAL SYSTEMS FOR 95 % (V/V) METHANOL IN WATER AS A SOLVENT

The counter ions were at their respective *pK** values. Driving current: 25 μ A in a 0.2 mm I.D. capillary

Parameter	<i>pH</i> * 5.84	<i>pH</i> * 6.82
Leading ion	Chloride	Chloride
Concentration	0.01 <i>M</i>	0.01 <i>M</i>
Counter ion	Hydroxylamine	Triethanolamine
Terminator	HEPES	BICINE*
Concentration	<i>ca.</i> 0.005 <i>M</i>	<i>ca.</i> 0.005 <i>M</i>

* BICINE = N,N-bis(2-hydroxyethyl)glycine.

ylamine or triethanolamine at their respective pK^* values. The operational systems used are listed in Table I.

The pH measurements were carried out with a Type 135 digital pH meter (Corning, Halstead, Essex, Great Britain), calibrated with aqueous standard buffer solutions (pH 6.0 and 8.0) (Merck, Darmstadt, G.F.R.). The pH values of the methanolic systems in Table I were not corrected for the methanol content and are therefore denoted with an asterisk. The pK^* values of the counter ions were determined by pH^* measurement of a solution of the buffer chloride salt, half of which was subjected to anion exchange. As the terminator for methanolic systems lithocholic acid has been used¹², but we prefer to use weak ions such as 2-morpholinoethanesulphonic acid (MES), N-2-hydroxyethylpiperazine-N'-2-ethanesulphonic acid (HEPES) and tris(hydroxymethyl)methylaminopropane sulphonic acid (TAPS). As in aqueous isotachopheresis, additives to the leading electrolyte may be needed to decrease the effect of electroosmosis¹², although these disturbances are expected to be of less importance in methanolic systems¹⁸.

The sample pre-treatment procedure used in the present investigation consisted of C-18 silica cartridge adsorption (Sep-Pak; Waters Assoc., Milford, MA, U.S.A.) as described by other workers⁹ for an HPLC method. The procedure was slightly modified. A volume of 20 μ l of human bile was diluted in 4 ml of 0.5 M phosphate buffer (pH 7.0) and applied to the cartridge, which was then washed with 15 ml of water and eluted with 3 ml of methanol. The average recovery of bile acids added to human bile was 91%. Microlitre amounts of the eluate or a standard solution of bile acid conjugates in methanol were injected directly. The time of analysis was 15–20 min at 25 μ A in a PTFE capillary of I.D. 0.2 mm. This longer than usual analysis time was caused by the relatively low transport number (<0.5) of the leading ion in the solvent used.

RESULTS AND DISCUSSION

One of the most important parameters used to achieve separation in isotachophoresis

Fig. 1. Isotachopheretic separation of five bile acids in a standard mixture. The operational conditions are summarized in Table I. The pH^* of the leading electrolyte was 6.82. 1 = 0.2 nmol taurochenodeoxycholic acid; 2 = 0.2 nmol taurocholic acid; 3 = 0.1 nmol glycochenodeoxycholic acid; 4 = 0.3 nmol glycocholic acid; 5 = 3.0 nmol cholic acid. R = Increasing resistance, t = increasing time.

phoresis, besides the difference in absolute mobility, is the difference in the pK values of the compounds to be separated. pK^* values for bile acids in methanol are not known, but pK values in water have been mentioned in the literature⁵: for free bile acids $pK = 5-6$, for glycine conjugates *ca.* 4 and for taurine conjugates *ca.* 2 was reported.

Fig. 1 shows the isotachophoretic separation at $pH^* 6.82$ of five bile acids in a standard mixture. From the fact that the non-conjugated cholic acid migrates with a considerably lower effective mobility than the conjugated bile acids, it is concluded that the pK^* value of cholic acid shows a positive shift with respect to its value in

Fig. 2. Isotachopheric separation of tauro- and glyco-conjugated acids according to their respective pK^* values at $pH^* 5.84$. The operational conditions are summarized in Table I. Zone numbers as in Fig. 1.

water. The same applies to the glyco-conjugated bile acids, as is illustrated in Fig. 2. Group separation on glyco- and tauro-conjugates is achieved at $pH^* 5.84$ so that the pK^* values of the glyco-conjugates can be calculated to be greater than 5. The pK^* of the tauro-conjugates remains below 5, as the effective mobility does not change in the interval $5.84 < pH^* < 6.82$.

These observations agree with reversed-phase liquid chromatographic data^{5,6} where at $pH^* 4.7$ the separation of the tauro- from the glyco-conjugates is mainly

Fig. 3. Analysis of a sample of human bile of a patient. Zone numbers as in Fig. 1. The concentrations of the individual bile acid conjugates were in the range 10–40 nmol/ μ l. A = Increasing UV absorption.

ascribed to pK^* differences. In reversed-phase chromatography there is also a significant effect of the number and/or orientation of the hydroxyl groups on retention behaviour⁶ in 65% methanol. In isotachopheresis in 95% methanol these effects were less pronounced, but sufficient to achieve complete separation of conjugated cholic and chenodeoxycholic acids, as illustrated in Fig. 2.

Fig. 3 shows the analysis of a sample of human bile of a patient (a suspected bile stone former). Taurochenodeoxycholic, taurocholic, glycochenodeoxycholic and glycocholic acids were determined. Because of the group separation at this pH^* , the clinically significant glyco-tauro ratio is easily seen.

We are currently investigating the possibilities of adapting the method for the determination of bile acids at serum levels. One of the advantages of using a more hydrophobic solvent in isotachopheresis is that a number of sample pre-treatment procedures used in chromatography can be applied to remove unwanted polar substances such as sodium chloride.

This study has shown that the use of non-aqueous isotachopheresis is very promising. For the determination of bile acid conjugates it compares favourably with liquid chromatography in terms of sensitivity and analysis time.

ACKNOWLEDGEMENTS

The human bile samples were kindly supplied by A. Roovers, Streeklaboratorium "Zeeland", Goes, and Dr. A. W. Mulder, Catharinaziekenhuis, Eindhoven, The Netherlands.

REFERENCES

- 1 M. T. Ravi Subbiah, *Ann. Clin. Lab. Sci.*, 3 (1973) 362.
- 2 H. Kimura, N. Suzuki, T. Sato, J. Goto and T. Nambara, *Jap. J. Clin. Chem.*, 8 (1979) 126.
- 3 G. Szepesi, K. Dudás, A. Pap, Z. Végh, E. Mincsovcics and T. Tyihák, *J. Chromatogr.*, 237 (1982) 137.
- 4 G. Lepage, C. C. Roy and A. M. Weber, *J. Lipid Res.*, 22 (1981) 705.
- 5 W. H. Elliott and R. Shaw, *Chromatogr. Sci.*, 16 (1981) 1-40.
- 6 C. A. Bloch and J. B. Watkins, *J. Lipid Res.*, 19 (1978) 510.
- 7 J. Goto, M. Hasegawa, H. Kato and T. Nambara, *Clin. Chim. Acta*, 87 (1978) 141.
- 8 T. Nambara, J. Goto, M. Hasegawa and H. Kato, in G. L. Hawk (Editor), *Biological and Biomedical Applications of Liquid Chromatography II*, Marcel Dekker, New York, 1978.
- 9 A. Th. Ruben and G. P. van Berge-Henegouwen, *Clin. Chim. Acta*, 119 (1982) 41.
- 10 S. Kamada, M. Maeda and A. Tsuji, *J. Chromatogr.*, 272 (1983) 29-41.
- 11 S. Baba, R. Uenoyama, K. Suminoe, F. Takeda, S. Hasegawa and Y. Kamenno, *Kobe J. Med. Sci.*, 26 (1980) 89.
- 12 F. M. Everaerts, J. L. Beckers and Th. P. E. M. Verheggen, *Isotachopheresis—Theory, Instrumentation and Applications (J. Chromatography Library, Vol. 6)*, Elsevier, Amsterdam, 1976.
- 13 A. Baldesten and S.-G. Hjalmarsson, *CRC Crit. Rev. Anal. Chem.*, 11 (1981) 261-352.
- 14 Z. Deyl (Editor), *Electrophoresis (J. Chromatography Library, Vols. 18A and B)*, Elsevier, Amsterdam, 1982.
- 15 J. L. Beckers, *Thesis*, Eindhoven University of Technology, Eindhoven, 1976.
- 16 D. Kaniansky, *Thesis*, Komensky University, Bratislava (1982).
- 17 Th. P. E. M. Verheggen and F. M. Everaerts, *J. Chromatogr.*, 249 (1982) 221-230.
- 18 J. C. Reijenga, G. V. A. Aben, Th. P. E. M. Verheggen and F. M. Everaerts, *J. Chromatogr.*, 260 (1983) 241-254.