Comment on "On the spatial and temporal resolution of ion temperature measurements based on neutral beam scattering"

Citation for published version (APA):

Heesch, van, E. J. M. (1987). Comment on "On the spatial and temporal resolution of ion temperature measurements based on neutral beam scattering". *Review of Scientific Instruments*, 58(7), 1322-1322. <https://doi.org/10.1063/1.1139413>

DOI:

[10.1063/1.1139413](https://doi.org/10.1063/1.1139413)

Document status and date:

Published: 01/01/1987

Document Version:

Publisher's PDF, also known as Version of Record (includes final page, issue and volume numbers)

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.tue.nl/taverne

Take down policy

If you believe that this document breaches copyright please contact us at:

openaccess@tue.nl

providing details and we will investigate your claim.

Comment on "On the spatial and temporal resolution of ion temperature measurements based on neutral beam scattering" [Rev. Sci. Instrum. 58, 1320 (1987)]

E. J. M. van Heesch

Eindhoven University of Technology, Department of Electrical Engineering, P. O. Box 513, 5600 MB Eindhoven, The Netherlands

(Received 10 March 1987; accepted for publication 18 March 1987)

The discrepancies found by Donné and Barbian [A. J. H. Donné and E. P. Barbian, Rev. Sci. Instrum. 58, 1320 (1987)] can be attributed principally to the choice of experimental and technical limits, basic assumptions, and definitions.

Independently performed calculations are the basis of our publication on the design of an ion-temperature diagnostic for STOR-M (hereafter called Ref. 1). We clearly notice the different results found by Donné and Barbian described in their letter (Ref. 2). In our view, these discrepancies can be attributed principally to the choice of experimental and technical limits, basic assumptions, and definitions.

To demonstrate the importance of these choices, we describe a few details distinguishing our concept from the approach in Ref. 2.

Since the total of additional broadening $\Sigma\sigma_i^2$, including the effect of angular spread, can be determined separately in our setup, its contribution to s^2 does not necessarily enter as a part of $\Delta T_i/T_i$.¹ In Ref. 1, for example, the relative corrections on s^2 are $\sigma_E^2/\sigma^2 = 0.01$ and $\sigma_\theta^2/\sigma^2 = 0.10$. The total number of 200 counts per spectrum, therefore, is not affected and remains an acceptable goal to achieve 10% accuracy.

Our design value for the time resolution is based on this number and on the high counting performance achievable in the scheme of MCP and adjacent local electronics.

This, basically, is not compatible with a TOF analyzer using a single-output, small-area MCP.

With respect to the spatial resolution, one should distinguish the directions parallel and perpendicular to the length l of the scattering volume. For STOR-M the perpendicular direction is chosen to be radial in the torus midplane.¹ The respective relevant spatial resolution is equal to the beam diameter of 1 cm. For the other direction (poloidal in STOR-M) one can use $\delta x = 1/3 l$ (2.2 cm), since the underlying assumption is not too unrealistic along a line piece l tangential to the equilibrium fluxsurface.

We consider the discussed scattering experiment as a very promising diagnostic tool. It is anticipated that the large amount of valuable data attainable justifies the effort needed to fully exploit this experiment.

¹E. J. M. van Heesch, A. Hirose, A. Sarkissian, and H. M. Skarsgard, Rev. Sci. Instrum. 57, 1792 (1986).

²A. J. H. Donné and E. P. Barbian, Rev. Sci. Instrum. 58, 1320 (1987).