Computer-aided diagnosis in chest radiography : a survey

Citation for published version (APA):

Ginneken, van, B., Haar Romenij, ter, B. M., & Viergever, M. A. (2001). Computer-aided diagnosis in chest radiography : a survey. *IEEE Transactions on Medical Imaging*, 20(12), 1228-1241.
<https://doi.org/10.1109/42.974918>

DOI:

[10.1109/42.974918](https://doi.org/10.1109/42.974918)

Document status and date:

Published: 01/01/2001

Document Version:

Publisher's PDF, also known as Version of Record (includes final page, issue and volume numbers)

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.tue.nl/taverne

Take down policy

If you believe that this document breaches copyright please contact us at:

openaccess@tue.nl

providing details and we will investigate your claim.

Computer-Aided Diagnosis in Chest Radiography: A Survey

Bram van Ginneken*, Bart M. ter Haar Romeny, and Max A. Viergever, Member IEEE

Abstract—The traditional chest radiograph is still ubiquitous in clinical practice, and will likely remain so for quite some time. Yet, its interpretation is notoriously difficult. This explains the continued interest in computer-aided diagnosis for chest radiography. The purpose of this survey is to categorize and briefly review the literature on computer analysis of chest images, which comprises over 150 papers published in the last 30 years. Remaining challenges are indicated and some directions for future research are given.

Index Terms—Chest radiography, computer-aided diagnosis, lung field segmentation, lung nodule detection, rib segmentation, .

I. INTRODUCTION

THE discovery of X-rays by Wilhelm Conrad Röntgen, in 1895 [1], has revolutionized the field of diagnostic medicine. Today a wide variety of (three-dimensional) imaging techniques is available, and many types of examinations have been, or are about to be replaced by computed tomography (CT) and/or magnetic resonance imaging (MRI) within the next few years. But this is certainly not the case for the chest radiograph. On the contrary, the traditional chest study is still by far the most common type of radiological procedure, making up at least a third of all exams in a typical radiology department [2], [?]. Daffner [2] describes the chest as “the mirror of health and disease.” An enormous amount of information about the condition of the patient can be extracted from a chest film and, therefore, “the ‘routine’ chest radiograph should not be considered quite so routine.”

Interpreting a chest radiograph is extremely challenging. Superimposed anatomical structures make the image complicated. Even experienced radiologists have trouble distinguishing infiltrates from the normal pattern of branching blood vessels in the lung fields, or detecting subtle nodules that indicate lung cancer [4]. When radiologists rate the severity of abnormal findings, large interobserver and even intraobserver differences occur [5], [6]. The clinical importance of chest radiographs, combined with their complicated nature, explains the interest to develop computer algorithms to assist radiologists in reading chest im-

ages. The purpose of this survey is to categorize and briefly review the literature on computer analysis of chest radiographs with an emphasis on the techniques that have been employed and on the tasks these techniques are supposed to solve.

Soon after the invention of the modern digital computer at the end of the 1940s, research began on having computers perform tasks that had previously been performed only by human intelligence. The first articles about computer analysis of radiographic images appeared in the 1960s [7], [8]. Papers describing techniques specifically designed for computerized detection of abnormalities in chest radiographs began to appear in the 1970s. These were among the first papers in the field of medical image analysis, as is noted in a recent review by Duncan and Ayache [9].

These early studies displayed a considerable optimism regarding the capabilities of computers to generate complete diagnoses. They are summarized in one review [10] as attempts to “fully automate the chest exam.” Over the decades this expectation has subsided (which seems to have happened to the early enthusiasm regarding the capabilities of artificial intelligence systems in general). Currently, the general agreement is that the focus should be on making useful computer-generated information available to physicians for decision support rather than trying to make a computer act like a diagnostician [11]; from the abbreviations FACD (fully automatic computer diagnosis), ICD (interactive computer diagnosis), and CAD (computer-aided diagnosis), only the latter is in common use nowadays.

Several related fields are important for CAD in chest radiography, but are outside the scope of this survey because no image processing is involved. Among these are studies on acquisition, about digital chest units versus analog film systems (e.g., [12]–[15]), or on dual energy systems in chest radiography (e.g., [16]–[18]); studies that use psychophysics, e.g., to determine the optimum tube voltage [19] or to aid the detection of abnormalities by measuring visual dwell [20], [21]; attempts to quantify the performance of radiologists, usually through observer studies, when image quality parameters of the chest radiographs are varied (e.g., [22]–[28]); and research on estimating probabilities that patients exhibit a certain disease, given a number of features from clinical information and/or output of computer algorithms. Such studies have continued to appear, starting in the 1960s [29], [30] until recently, e.g., [31]–[33].

Three main areas can be distinguished in the literature on computer analysis of chest radiographs: 1) general processing techniques; 2) algorithms for segmentation of anatomical structures; and 3) analysis aimed at solving a particular task or ap-

Manuscript received June 10, 2001; revised October 30, 2001. Asterisk indicates corresponding author.

*B. van Ginneken is with the Image Sciences Institute, University Medical Center Utrecht, Heidelberglaan 100, 3584 CX Utrecht, The Netherlands (e-mail: bram@isi.uu.nl).

B. M. ter Haar Romeny is with the Faculty of Biomedical Engineering, Medical and Biomedical Imaging, Eindhoven University of Technology, 5600 MB Eindhoven, The Netherlands.

M. A. Viergever is with the Image Sciences Institute, University Medical Center Utrecht, Heidelberglaan 100, 3584 CX Utrecht, The Netherlands.

Publisher Item Identifier S 0278-0062(01)11136-5.

plication, usually an attempt to detect a specific kind of abnormality. The following subdivision is adopted and followed in the sequel of this survey.

- General processing:
 - enhancement;
 - subtraction techniques.
- Segmentation:
 - lung fields;
 - rib cage;
 - other structures.
- Analysis:
 - size measurements;
 - lung nodule detection;
 - texture analysis;
 - other applications.

Connors *et al.* [10] reviewed computer analysis of radiographic images in general in 1982, with a strong emphasis on chest radiography. Overviews of research at the Kurt Rossmann Laboratories at the University of Chicago, an active group in computer-aided diagnosis, are given in [34]–[40]. A recent general introduction to computer-aided diagnosis with some examples taken from CAD in chest radiography can be found in [41]. Other informative general overviews can be found in [42] and [43], although the discussion on CAD in chest radiography in both works is strongly focussed on the work done at the University of Chicago. An interesting collection of short papers on CAD can be found in [44]. Reeves and Kostis [45] recently reviewed CAD for lung cancer and although the emphasis is on CT, there is also an overview of methods developed for chest radiographs.

It is worth mentioning here that CAD has been shown to be effective in screening mammography. These successes are certainly helping to make CAD in chest radiography acceptable, notably for the detection of lung nodules. Moreover, the techniques employed in both fields are often similar and may inspire new research. Overviews of CAD in mammography can be found in [42], [46], and [47].

Almost all work discussed here is applied to frontal [posterior-anterior (PA)] chest radiographs. A few studies are aimed at lateral radiographs, and those cases are specifically mentioned.

Although the skills of experienced radiologists are beyond reach for nonmedical researchers, anyone working on algorithms for the analysis of chest radiographs should obtain some knowledge about the anatomy of the chest, its appearance on projection radiographs and the nature of various abnormal findings. Introductory texts can be found in Squire [48], and in Lange's standard textbook [49]. A detailed pictorial description of normal and abnormal findings in chest radiographs is given in the chapter on chest imaging in [50]. A comprehensible overview of interstitial disease can be found in [51].

Several tables are included that summarize papers on specific topics. In these tables, studies are listed in chronological order. In most cases information about the evaluation of the presented techniques, such as the size of the test database, are included. We refrained from including any reported figures on classification performance or segmentation accuracy, because these results are usually very dependent on the subtlety of the abnormalities in the test images which makes it impossible to compare the merits of various methods on the basis of such figures alone.

II. GENERAL PROCESSING

A. Enhancement

Chest radiographs inherently display a wide dynamic range of X-ray intensities. In conventional, unprocessed images it is often hard to “see through” the mediastinum and contrast in the lung fields is limited. A classical solution to this kind of problem in image processing is the use of (local) histogram equalization techniques. A related technique is enhancement of high-frequency details (sharpening). Such techniques are so essential for optimized display of images in soft-copy reading environments that numerous studies have been devoted to this subject. A wide variety of preprocessing procedures for chest radiographs based on local equalization, sharpening, and combinations and modifications of these techniques have been proposed. It is beyond the scope of this survey to give a complete overview. The following works deal with optimal display and image enhancement [52]–[61]. Examples of hardware solutions can be found in [62], [63]. Nowadays, almost all vendors sell digital chest units with a larger dynamic range than conventional units and most vendors automatically preprocess the images with proprietary algorithms.

B. Subtraction Techniques

Subtraction techniques attempt to remove normal structures in chest radiographs, so that abnormalities stand out more clearly, either for the radiologist to see or for the computer to detect.

A first approach is temporal subtraction, proposed by Kano *et al.* [64]. An input image is registered with a previous radiograph of the same patient. Elastic matching is employed in which the displacement of small regions of interest (ROIs) is computed based on cross correlation and a smooth deformation field is obtained by fitting a high-order polynomial function to the displacement vectors. The registered image is subtracted and if the registration is successful, areas with interval change stand out as either dark or bright on a gray background. The original technique has been improved and evaluated using subjective ratings by radiologists [65], [66] and the results of the method have been compared with manual registration [67]. An observer study [68] with a small number of selected cases showed an increase in detection accuracy of interval change when both the normal image and the subtraction image were presented to the radiologist. Recently, Zhao *et al.* [69] proposed a temporal subtraction technique that performs elastic registration on rib border segments detected with an adaptive oriented filter and reported that the contrast around lung nodules substantially increased in the subtracted images. With the advent of digital archives, temporal subtraction techniques may be applied on a routine basis since monitoring interval change is one of the main reasons for making chest radiographs (it is common to obtain a chest radiograph of patients on an intensive-care unit every day).

If a previous radiograph is not available, a subtraction can be made by mirroring the left/right lung field, performing elastic registration on the right/left lung field and subtracting. This technique, coined *contralateral subtraction*, employs the symmetry of the rib cage. Li *et al.* [70], [71] use a scheme similar to Kano *et al.* [64], based on cross correlation and evaluate the

TABLE I

AN OVERVIEW OF LITERATURE ON LUNG FIELD SEGMENTATION. THE FIRST COLUMN LISTS THE FIRST AUTHOR AND REFERENCE(S). METHODS ARE DIVIDED IN RULE-BASED (RB) AND/OR PIXEL CLASSIFICATION (PC) SCHEMES, SEE THE TEXT FOR DETAILS. THE COLUMN EVALUATION GIVES THE NUMBER OF IMAGES USED IN EVALUATION ONLY (IMAGES USED IN A TEST SET ARE NOT INCLUDED). R MEANS EVALUATION THROUGH SUBJECTIVE RATING BY A RADIOLOGIST OF THE RESULT, USUALLY IN CLASSES FROM 1 (POOR) TO 5 (EXCELLENT). Q STANDS FOR QUANTITATIVE EVALUATION, IN WHICH THE RESULT IS COMPARED ON A PIXEL-BY-PIXEL BASIS WITH A MANUAL SEGMENTATION

	RB	PC	PA/lateral	evaluation	remarks
Toriwaki [77], [78]	×		PA	none	describes a complete analysis system
Harlow [79]	×		PA	none	
Chien [80], [81]	×		PA	none	only right lung, used to detect abnormalities
Hasegawa [82]		×	PA	none	
Pietka [83]	×		PA	none	
McNitt-Gray [84], [85]		×	PA	16 Q	uses 5 anatomical classes
Duryea [86]	×		PA	802 Q	
Xu [87]	×		PA	1000 R	outer rib cage only
Xu [88]	×		PA	300 R	diaphragm edges only
Armato [89]	×		PA	600 R	costophrenic angles only
Armato [90]	×		lateral	200 Q	
Armato [91]	×		PA	600 R	
Carrascal [92]	×		both	65 RQ	
Vittitoe [93]		×	PA	99 Q	
Tsujii [94]		×	PA	71 Q	
Wilson, Brown [95], [96]	×		PA	none	describes a complete analysis system
Vittitoe [97]		×	PA	115 Q	uses 6 anatomical classes
Van Ginneken [98]	×	×	PA	115 Q	
Van Ginneken [99]			PA	115 Q	general knowledge-based method

result with subjective ratings by radiologists. Yoshida [72], [73] minimizes the sum of squared differences, using a mapping that is ensured to be smooth through a regularization term and an optimization in the wavelet domain, and, in [74] segments abnormal areas that “light up” in the subtracted image. The technique is used to eliminate false positives from a number of candidate lung nodules.

It may also be possible to eliminate or suppress normal structures in radiographs by subtracting a model that is fitted to the input image. The algorithm to remove ribs by Vogelsang *et al.* [75], [76] is an example of this approach. Ribs borders are detected and a simple physical model of the density of the rib cage is subtracted from the input image. The technique is not evaluated.

III. SEGMENTATION

A. Lung Fields

Automatic segmentation of the lung fields is virtually mandatory before computer analysis of chest radiographs can take place. Several studies deal with this problem exclusively. Others deal with parts of the problem, such as detection of the outer ribcage [87], the diaphragm [88], or the costophrenic angle (where the diaphragm and the rib cage meet) [89]. A few studies focus on lateral chest radiographs [90], [92].

Table I gives an overview of papers on lung segmentation. The two main approaches are rule-based reasoning and pixel classification. A rule-based scheme is a sequence of steps, tests and rules. Most algorithms for the segmentation of lung fields fall in this category [83], [86]–[92], [98]. Techniques employed are (local) thresholding, region growing, edge detection, ridge detection, morphological operations, fitting of geometrical models or functions, dynamic programming. On the other hand, several attempts have been made to classify each pixel in the image into an

anatomical class (usually lung or background, but in some cases more classes such as heart, mediastinum, and diaphragm [85], [97]). Classifiers are various types of neural networks, or Markov random field modeling, trained with a variety of (local) features including intensity, location, texture measures [84], [85], [93], [94], [97]. Van Ginneken and Ter Haar Romeny [98] combine both approaches in a hybrid scheme.

It is also possible to use general knowledge-based segmentation methods, such as active shape models [100], or extensions of such methods for the segmentation of lung fields, as is shown in [99].

In studies done in the 1970s often a rough or partial outline of the lung fields was detected, with rule-based schemes usually based on analysis of profiles [101]–[105]. In this survey, the term “profiles” refers to an average of consecutive one-dimensional lines of pixels, usually running horizontally or vertically. Evaluations of the effectiveness of these methods are not made, and none of these papers focuses on lung field segmentation specifically but use the segmentation for further processing or to estimate the heart size or lung capacity. These methods undoubtedly inspired later studies.

Overall, the problem of segmenting lung fields has attracted considerable attention. Several authors report good results that approach the inter-observer variability ([98] lists a comparison of quantitative results of several studies). The task of segmenting lung fields might, therefore, be considered largely solved, although no attempts have been made to test methods with very large databases to verify if schemes are also able to produce reasonable results for the, say, 1% most difficult cases. Likewise, a study in which the performance of various methods on a large common database of radiographs (of varying quality and obtained with different settings) is compared, has not been made, and would be a worthwhile endeavor.

TABLE II

AN OVERVIEW OF LITERATURE ON RIB DETECTION IN FRONTAL CHEST RADIOGRAPHS. THE FIRST COLUMN LISTS THE FIRST AUTHOR AND REFERENCE(S). THE COLUMN P/A INDICATES IF POSTERIOR (P) OR ANTERIOR (A) RIBS ARE DETECTED, OR BOTH OR NONE. THE RIGHT COLUMN LISTS ON HOW MANY IMAGES THE METHOD IS EVALUATED AND IF THE EVALUATION IS A SUBJECTIVE JUDGEMENT FROM THE AUTHOR (A), BASED ON RATINGS BY RADIOLOGISTS (R) OR QUANTITATIVE (Q)

	geometric model	detection method	P/A	evaluation
Toriwaki [77], [78]	parabolas	edge detection, parabolas fitted through candidates, candidates are merged	PA	15A, 40A
Wechsler [107], [108], [109]	parabolas, ellipses, 4th order polynomials	oriented edge detection, modified Hough transform and rules to link segments	PA	5A
Ballard [110]	–	models relations between ribs to classify rib border candidates	P	4A
De Souza [111]	a combination of two parabolas	vertical edge detection, grouping, rejecting, refining candidates and fitting	PA	8A
Powell [112]	detects small square ROIs in inter-rib spaces	fits a periodic function with varying frequency for costal and intercostal space to pixel data	–	66Q
Sanada [113]	ellipses with height given by function from [112]	vertical edge detection, grouping candidates and fitting geometric model	P	50R
Yue [114]	parabolas, snakes	edge detection, modified Hough transform, segment selection rules, followed by snake fitting	P	10Q
Vogelsang [75], [76]	parabolas, snakes, active shape models	edge detection, template matching, modified Hough transform, rules to select among candidates	P	154A
Van Ginneken [115], [99]	parabolas, statistical rib cage model from principal component analysis	optimization of model parameters with cost function based on oriented edge detection	P	35Q

Finally, we note that it is often hard to establish ground truth data for evaluation of segmentation techniques. This problem applies to all segmentation tasks discussed here. Preferably, the objects should be delineated by several experienced observers. In the case of lung fields, the definition of a lung field is already unclear, because large parts of the lungs are obscured in PA chest radiographs [106]. In rib segmentation, the overlapping posterior and anterior rib parts make a definition of “rib borders” difficult.

B. Rib Cage

There are several reasons why the automatic delineation of rib borders can be useful for computer analysis. First of all, the ribs provide a frame of reference and the locations of abnormalities are often indicated by radiologists in terms of numbered (inter) costal spaces. Second, rib border segmentation may be used to detect rib abnormalities such as fractured or missing ribs. Third, once locations of ribs are known, ROIs between (the intercostal space) or on (the costal space) the ribs can be defined for further analysis. Fourth, knowledge about the location of rib borders may be used to eliminate false positives in the detection of abnormalities such as nodules. Rib crossings (of the posterior and anterior parts of the ribs) frequently turn up as false positives in nodule detection schemes.

The approaches to rib (cage) segmentation are summarized in Table II. Several methods also detect (parts of) the contours of the lung fields [111], [112], [114], but an evaluation of their accuracy is not included. The classical approach to rib segmentation, used in most studies, is as follows. A geometrical model of rib borders

is chosen, e.g., parabolas or ellipses or a combination. Edge detection extracts segments that may be parts of the rib borders. These candidate rib border parts are analyzed and rejected, or grouped together into complete ribs. In other studies, each candidate votes for combinations of model parameters (a modified Hough transform). Postprocessing steps may remove some borders, or infer new rib borders that had not been detected. A fine-tuning stage may be added in which the strict geometric model is abandoned, such as the snakes fitted to rib borders in [114].

A difficulty in the classical approach is that the relations between ribs are not taken into account during the fitting procedure. Ribs may be missed or detected twice and the fact that consecutive ribs and left-right ribs have similar shape is often ignored. Early attempts to model the relation between ribs, as in [110] do not appear very powerful or use simple *ad hoc* rules [114]. This was already recognized by Wechsler [109] in 1977: “We believe that at this stage a significant improvement in rib boundary detection could be achieved only by a knowledge-based system which would incorporate a model of the rib cage.” The rib-cage model in [115] is an attempt to model the shape of the complete rib cage and fit it directly to the image.

The number of images used in evaluation is in general considerably smaller than in the case of lung field segmentation and in many cases the evaluation is confined to qualitative judgements from the authors themselves (these studies are indicated with A in the right column of Table II). Since the full rib cage has a much more complex shape than the outline of the lung fields, partial failures occur frequently. Furthermore, the detection of anterior (or ventral, as opposed to the posterior or dorsal) parts of the ribs has been addressed only in studies that omit proper

evaluation. Therefore, we conclude that the problem of automatic segmentation of the rib cage in chest radiographs is still far from solved.

C. Other Structures

Toriwaki *et al.* describe a complete system for the analysis of chest radiographs [77], [78] that includes segmentation of lung fields, rib cage, heart, clavicles, and blood vessels followed by automatic detection of abnormalities. Only a rough description of results on 15 [77] and 40 images [78] taken from a mass screening for lung tuberculosis in Japan is given without mention of the nature and subtlety of the abnormalities in this database.

A method to select small ROIs within the lung fields that do not contain rib borders is presented by Chen *et al.* [116]. The lateral lung fields are subdivided into regions which are partly eliminated if they contain edges of certain strength and orientation.

Several methods for segmentation of (parts of) the heart boundary have been proposed [8], [76], [101], [117]–[120], usually with the aim of detecting cardiomegaly (enlarged heart size). Note that a correct segmentation of the lung fields is sufficient to compute the cardiothoracic ratio (CTR) indicative of cardiomegaly, since parts of the boundaries of the lung fields coincide with the heart contour. In Section IV-A, these methods will be discussed in some more detail.

Armato *et al.* [121] developed an algorithm to detect abnormal asymmetry in chest radiographs using a rule-based scheme for detecting lung contours and comparing the projected lung areas. The method was evaluated on 70 radiographs.

IV. ANALYSIS

What kind of abnormalities does the radiologist who reads chest radiographs encounter in his/her day-to-day practice? The answer to this question should set the agenda for computer scientists who want to—as it was put in Section I—“fully automate the chest exam.”. MacMahon [122] has attempted to answer this question by counting the abnormalities encountered in 1085 abnormal chest radiographs, collected from consecutive cases, and dividing them in 30 categories. The main results of this study are summarized in Table III. The statistics are surprising when compared with abnormalities commonly investigated in the literature on computer analysis of chest images. Lung nodules are a relatively rare abnormality but have received much attention in the literature and observer studies tend to use their detectability as a criterion for diagnostic accuracy. Pulmonary infiltrates could be detected in principle by the texture analysis schemes to be discussed later and are the most commonly occurring abnormality. Catheters are next, and although catheters usually stand out clearly in the image, locating their tip was the task that was most frequently limited by image quality. One may argue that the presence of a catheter, or drainage tube, or pacemaker is not an abnormality. However, the presence of such objects can have a large effect on computer analyses and it is important to realize how common such findings are. The same is true for clothing artifacts, that are present in many chest radiographs.

The following applications of computer analysis of chest radiographs are reported in the literature.

- Lung nodule detection.

TABLE III

THE FREQUENCY WITH WHICH ABNORMALITIES ARE ENCOUNTERED IN ABNORMAL CHEST RADIOGRAPHS ENCOUNTERED IN CLINICAL PRACTICE. DATA WAS OBTAINED FROM 1085 STANDARD RADIOGRAPHS OF ADULTS. THE INFORMATION IN THIS TABLE IS TAKEN FROM THE STUDY “THE NATURE AND SUBTLETY OF ABNORMAL FINDINGS IN CHEST RADIOGRAPHS” BY MACMAHON *et al.* [122]

Finding	% of all abnormalities
Pulmonary infiltrates	55%
Intravenous catheters	33%
Heart size/contour	27%
Endotracheal/tracheostomy tubes	22%
Pleural effusions	12%
Linear atelectasis/scar	10%
Drainage catheters and tubes	9%
Pulmonary vascularity	9%
Pleural scarring	8%
Rib lesions	7%
Mediastinal masses	6%
Diaphragm	5%
Calcified granulomas	5%
Pneumothorax	5%
Lung nodules	5%
Extrathoracic abnormalities	4%
Lung masses	2%
Calcified nodules	1%
Mediastinal shift/contour	1%
Cardiac pacemakers	1%

- Detection of cardiomegaly or estimation of the CTR.
- Estimation of total lung volume.
- Detection of pneumothorax.
- Estimation of the severity of pneumoconiosis (coal worker's disease).
- Detection of interstitial disease.
- Detection of abnormalities encountered in mass screening for tuberculosis.

We divide these applications into four groups. Size measurements are applications based on results that can be computed directly from a segmentation. The detection of nodules is considered a separate category. Several other applications are grouped under “texture analysis” and the remaining tasks are listed under “other applications.”

A. Size Measurements

In a few applications, diagnostic information can be extracted directly from segmentations. An example is the estimation of enlarged heart size (cardiomegaly) usually performed by measuring the CTR, the maximal horizontal diameter of the heart divided by the maximal horizontal diameter of the thorax. This is the subject of probably the earliest computer analysis studies of chest radiographs by Becker, Meyers, and co-workers [7], [8]. Points on the rib-cage boundary and heart boundary are extracted from horizontal profiles and the CTR is estimated for 37 radiographs. In work by Hall [117] and Kruger [101], and their co-workers, several parts of the lung and heart boundaries were segmented. Images were classified in one of three categories (normal, rheumatic heart disease, or otherwise abnormal) using various ratios of distances derived from the segmentation as

TABLE IV
AN OVERVIEW OF METHODS FOR INITIAL SELECTION OF NODULE CANDIDATES. THE FIRST COLUMN LISTS THE FIRST AUTHOR AND REFERENCE(S). THE METHOD AND EVALUATION DATA SET ARE DESCRIBED BRIEFLY

study	method	evaluation
Sklansky [124], Ballard [125], [126]	Generalized Hough transform: edge detection, edges vote for centers of circles of different radii, using gradient magnitude. Non-maxima suppression.	6 images
Sklansky [127]	Spoke filter, comparable to [125], [126].	19 images, 16 abnormals
Lampeter [128]	A blurred version of the image is subtracted. The result is histogram equalized. Candidates are detected with a Hough transform, similar to [125].	37 images
Giger [129]	Subtraction image is produced by filtering with a spherical kernel and subtracting a median filtered version. Nodules candidates are found by thresholding.	real and simulated nodules
Yoshimura [130]	Subtraction image is produced by a morphological open operation and subtracting a median filtered version. Compared and combined with [129].	60 images, 30 abnormals
Suzuki [131], [132]	A non-linear filter with three circular bands subdivided in segments produces output if the maximum segment values of each band decrease.	192 nodules
Lo [133]	A subtraction image is constructed similar to [129]. On the subtraction image, template matching with cross-correlation and sphere profiles is performed, followed by thresholding.	unclear
Yoshida [134]	Nodule enhanced image is produced by computing least asymmetric Daubechies wavelet transform and amplifying responses from intermediate levels before backtransformation. A subtraction image is produced as in [129]. Detected nodules are combined with those detected by [129].	100 images, 122 nodules
Mao [135]	Similar to [125]; for each pixel it is tested if pixels at various distances r display an edge pointed towards the circle.	140 generated nodule images
Wei [136]	A set of three filters measure if the gradient of edge pixels point to a single center. The edges are located within or on the diameter of a circle, and the diameter and support of these circles is varied.	120 images

features in linear and quadratic discriminant functions. The method was evaluated on 320 films, a surprisingly large number for studies from the 1970s. Using a similar approach, Sezaki and Ukena [118] computed the CTR by a scheme that detects the vertical boundary of the rib cage and the heart through analysis of horizontal profiles and the application of a few rules to correct failures. Roellinger *et al.* [123] used the method from [101] to detect points on the heart boundary to which a Fourier shape was fitted. The coefficients of the Fourier shape were used to classify the heart shape as normal or abnormal, giving reasonable results on a database of 481 images with 209 normal cases. A comparable method was proposed by Nakamori *et al.* [119]. A Fourier shape was fitted to points on the heart boundary found by edge detection and the results were used to detect cardiomegaly in [120] and evaluated on a database of 400 radiographs with 91 abnormals. In conclusion, computer estimation of the CTR has received considerable attention and promising results have been obtained. It is likely that modern segmentation methods will outperform the schemes used in most studies cited here since these are based on simple rules to detect only a limited number of landmark points in the image. Clinical applications of automatic detection of heart shape abnormalities seem feasible, but the problem has not attracted much attention the last decade.

A related example of size measurements is the determination of total lung capacity (TLC). In this case boundary points on both a PA and a lateral radiograph must be detected. The volume is estimated using an empiric formula that assumes simple shapes for the lungs and that has been shown to correlate

well with the true TLC. Paul *et al.* [103] described a method to automatically determine the TLC using profile analysis similar to Kruger *et al.* [101]. It was tested on 15 radiographs. Carrascal *et al.* [92] used a rule-based segmentation of PA and lateral lung fields to estimate the TLC for 65 radiographs.

B. Nodule Detection

Automatic detection of lung nodules is the most studied problem in computer analysis of chest radiographs. One in every 18 woman and every 12 men develop lung cancer, making it the leading cause of cancer deaths. Early detection of lung tumors (visible on the chest film as nodules) may increase the patient's chance of survival. But detecting nodules is a complicated task; see, e.g., [4]. In a lung cancer screening program for heavy smokers in which chest radiographs were taken every four months, it was shown that for 90% of peripheral lung cancers that were detected, nodules were visible on earlier radiographs, when these older images were checked retrospectively [159].

Nodules show up as relatively low-contrast white circular objects within the lung fields. The difficulty for CAD schemes is to distinguish true nodules from (overlapping) shadows from vessels and ribs.

Almost all methods rely on a two-step approach for nodule detection. In the first stage initial candidate nodules are detected. The second stage consists of eliminating as many false positive candidates as possible, without sacrificing too many true positives. Table IV lists methods for the detection of candidate nod-

TABLE V

AN OVERVIEW OF METHODS FOR CLASSIFYING NODULE CANDIDATES. THE FIRST COLUMN LISTS THE FIRST AUTHOR AND REFERENCE(S). THE SECOND COLUMN DESCRIBES THE COMPUTED FEATURES. THE THIRD COLUMN LISTS THE CLASSIFIER WHERE ANN STANDS FOR A STANDARD FEED-FORWARD ARTIFICIAL NEURAL NETWORK. THE RIGHT COLUMN DESCRIBES THE DATABASE USED FOR EVALUATION (THIS IS NOT NECESSARILY THE NUMBER OF IMAGES AND NODULES USED FOR VALIDATION ONLY)

study	features	classifier	evaluation
Ballard [125], [126]	Radius, contrast, distance from hilum, shape features.	nearest neighbor	6 images
Sankar [137]	Shape and contrast features from boundary obtained with dynamic programming.	thresholds	8 images, 30 nodules
Sklansky [127]	13 features based on histograms and output of filters applied to ROIs.	thresholds	19 images, 16 abnormalities
Lampeter [138], [128]	Rib test (presence of parallel edges), spatial location measures, Hough transform response, radius of circle, gray level features.	linear discriminant	37 abnormalities
Giger [129]	Circularity, size and growth rate of region grown area for various thresholds.	thresholds	real and simulated nodules
Giger [139]	Circularity of region grown area for various thresholds after morphological opening.	threshold	60 images, 33 nodules
Cox [140]	Haralick's [141] and Laws' [142] texture features obtained from ROIs with and without nodules.	ANN	ROIs from 5 images
Matsumoto [143]	Area after region growing around candidates with different thresholds, degree of circularity and irregularity of regions, magnitude and orientation of single scale gradient within ROIs after background subtraction.	thresholds	60 images, 30 abnormalities, 32 nodules
Suzuki [131], [132]	First a non-linear filter, similar to the nodule detector, detects other structures such as vessels, ribs. For remaining candidates, density and circularity features from concentric bands are computed.	thresholds, linear discriminant, feature selection	192 nodules
Peter Chiou [144]	Single scale edge strength and orientation.	Kohonen map	31 images, all abnormal, 87 nodules
Wu [145]	Similar to [143].	ANN, discriminant analysis, thresholds	60 images, 30 abnormal, 32 nodules
Lo [146]	32 × 32 raw image data from background corrected candidate ROIs.	convolution neural net	55 images, 25 abnormalities, 52 nodules
Lin [147], [148]	32 × 32 raw image data from background corrected candidate ROIs.	two convolution neural nets	In [148]: 54 images, 23 abnormalities
Floyd [149]	Fractal dimension estimated from power spectrum.	threshold	20 images, 10 abnormal
Vittitoe [150]	Fractal dimension estimated from power spectrum.	threshold	30 images, real and simulated nodules
Xu [151]	Diameter, circularity and irregularity measures of candidates obtained by region growing using several thresholds, the slopes of these measures, profile measures, size of regions as a function of thresholds, contrast and gradient measures.	thresholds followed by ANN	200 images, 100 abnormalities, 122 nodules
Carreira [152]	Size as a function of thresholds, various shape features from the intensity landscape.	thresholds	35 abnormal images, 70 nodules
Penedo [153]	Curvature of intensity landscape	ANN	60 images, 90 nodules, 288 simulated nodules
Xu [154]	Mean, minimum, maximum, width, standard deviation of histograms of gradients perpendicular to candidate boundaries for various segments of the boundary.	thresholds	200 images, 100 abnormalities, 122 nodules
Casaldi [155]	Morphological operations are used to detect likely location of nodules.	—	applied to lossy coding
Nakamura [156]	Computed from outlines of nodules manually determined by radiologists: mean, standard deviation, circularity, ellipticity, irregularity, root-mean-square variation, first moment power spectrum, tangential gradient index, radial gradient index, line enhancement index, mean gradient.	ANN classifies nodules as benign / malignant	56 cases, 34 malignant, 22 benign nodules
Yoshida [73]	After local contralateral subtraction to remove rib structures, signal to noise ratio in candidate ROIs.	threshold	550 candidates, 51 nodules
Catarious [157]	Channeled Hotelling statistics, fractal dimension, geometric features	ANN	237 candidates
Li [158]	Template matching with nodule and non-nodule candidates	threshold	228 images, 250 nodules

ules. Table V lists schemes for eliminating false positive candidates. Several methods include both steps but focus on one of them in particular. These methods are included in the most appropriate table.

Several schemes start with producing an image in which nodules are enhanced. This is done by filtering with a nodule-like filter and/or suppressing background structures by some sort of blurring [129], [133], or by applying preprocessing techniques such as unsharp masking or similar operations [138]. Nodule

candidates are detected using template matching or a modified Hough transform in which edge pixels vote for circles that could cause these edges [125], [126], [128], [135], [138]. In other cases the nodule enhanced image is simply thresholded. Some studies apply a background trend correction on ROIs with nodules candidates [143], [154], which is similar to high-pass filtering of the input image as a preprocessing step.

The next step consists of eliminating false positive responses. Usually specific features are detected for each nodule candi-

date. Table V gives a brief description of features used in each method. Almost all features are (variations on) “classical features” in the sense that they can be found in standard image processing textbooks, see for instance [160, Chapter 9]. Classifiers are remarkably often simple thresholds (sometimes of linear combinations of features) but classifiers from pattern recognition theory and standard feed-forward neural networks are also used. A few studies focus on the use of a particular type of weakly connected networks [146], [148].

It is clear from Table IV that there is considerable overlap between the methods employed in various studies (often by the same researchers; the number of groups that have been working on this problem is limited). Judging the relative merits of methods on the basis of these papers alone seems hardly possible. Comparisons between different methods are rare; sharing databases or setting up common databases for the evaluation of complete systems has not been given much attention. An exception, and probably the first attempt to set up a common database for lung nodules is the work by Shiraishi and co-workers [161], who collected a database of 247 chest radiographs with 154 nodules that is available on CD-ROM. The subtlety of the nodules in the database has a profound impact on detection accuracy and sensitivity, much more than in the case of lung field or even rib-cage segmentation. We believe that creating larger databases is crucial to progress in this field. Unless these databases are publicly available, it will remain difficult to compare performance of different methods (see also [162]). The evaluation is an important part of studies on computerized nodule detection. It is encouraging that receiver operating characteristic methodology [163], [164] has become the standard tool for evaluation. In general, nodule detection remains an open and difficult task in computer-aided diagnosis that is still far from solved.

A related problem is the classification of nodules as benign or malignant. Nakamura *et al.* [156] describe a method in which a feed-forward neural network outperformed radiologists in classifying nodules as benign/malignant based on features computed from manually drawn outlines of the nodules. Gurney and Swensen [165] used a Bayesian classifier and a feed-forward neural network to discriminate benign and malignant nodules using only manually determined features. They reported good results (and the Bayesian classifier outperformed the neural network). A preliminary conclusion, based on these two studies, could be that even if the performance of a computer in *detecting* nodules still lags that of radiologists, computer algorithms may serve a role in estimating the likelihood of malignancy once nodules are detected. Computers are possibly better than radiologists at extracting diagnostic information from large numbers of features, which include both (quantitated!) radiological findings and clinical data. It would be worthwhile to test CAD schemes for nodule classification on larger databases and to investigate if the manual steps in the methods described in [156] and [165] can be automated.

Not mentioned in Table IV are observer studies that test the application of CAD schemes for nodule detection [139], [166]–[168]. These studies are of utmost importance for testing the usefulness of nodule detection systems in practice. It is highly unlikely that nodules detection systems will be used

“stand-alone” without a radiologist. Therefore, the only way to assess the value of a CAD method is to compare the performance of radiologists with and without computer assistance. A CAD method may obtain good results in terms of high true positive detection rates and few false positives per image, but the actual value of the system may be much higher if it is likely to detect those nodules that are likely to be missed by radiologists. Therefore, it is also worthwhile to pay attention to studies that assess the capabilities of radiologist to detect nodules, such as [4], [169], and [170].

Methods for synthesis of images with simulated lung nodules can be useful to enlarge existing nodule databases. An example is the work of Sherrier *et al.* [171], who construct simulated nodules images by identifying isolated nodules, without any rib structures nearby, and adding rotated and scaled versions of these nodule templates to new radiographs, at various locations. Physically there is not much wrong with this procedure since radiographs are projection images, although it could be doubted whether the distribution of size and structure of nodules does not depend on the location in the lung field, and the scatter contributions of nodules of different size and orientation would be different. Observer studies in [171] showed that radiologists could not discern between simulated and true nodule cases.

C. Texture Analysis

Many diseases of the lungs are characterized by diffuse patterns in chest radiographs. These diseases are often referred to as diffuse lung disease or interstitial disease (although strictly speaking the latter is a smaller category). The interstitium of the lung is the connective tissue between the blood vessels and the alveoli, the tiny air sacs. Sorting out interstitial disease is one of the most difficult tasks for a chest radiologist [51]. Clues are the type of patterns (linear, reticular, nodular, honeycomb), the location, shape, and symmetry of the affected areas, the borders of these areas (well- or ill-defined) and their changes over time. From an image processing point of view, texture analysis is the proper way to analyze these kinds of abnormalities, hence, the title of this section.

The difference between computer detection of nodules and that of interstitial abnormalities is that interstitial abnormalities have a more diffuse character and, therefore, the two-stage approach for lung nodule detection (finding candidates and eliminating false positives) is not easy to apply. Instead, all areas within the lung fields should be checked for the presence of patterns of interstitial disease.

In the 1970s, texture analysis in chest radiographs was applied to the detection of pneumoconiosis, or coal miners’ disease. Most of this work used one or more of the by then state-of-the-art methods for texture feature generation, which are reviewed by Haralick [172]. Revesz and Kundel investigated the feasibility of classification based on features computed from the (optically determined!) Fourier spectrum [173]. A similar device was used later by Stark and Lee [174] who compared the performance of several classifiers with a database of 64 chest films. Kruger, Turner, and Thompson [175], [176] computed features derived from co-occurrence matrices (cf. [141]) and the Fourier spectrum of manually selected ROIs and

classified these regions with linear discriminant analysis. A similar approach was advocated by Sutton and Hall [102]. Later work by Hall [104], [177], used a coarse method for automatic segmentation of the lung fields and presented the design for a complete system. Ledley *et al.* [178] used texture features based on the size and shape of binary versions of the input image obtained by thresholding. They performed classification experiments with 64 films. Jagoe and Patton explored the use of features based on the magnitude and direction of the gradient in several studies [105], [179], [180]. Li *et al.* [181] treated the problem as one of detecting small rounded opacities and employed a strategy comparable with the standard nodule detection strategy. Classification is based on contrast features computed from region grown areas. Recently, Soliz *et al.* [182] presented a system for the detection of pneumoconiosis using a specific type of neural network and features derived from cooccurrence matrices from manually selected ROIs. To our knowledge, none of the work on detecting pneumoconiosis has been used in practice or evaluated for clinical use. This application would be ideal for testing techniques that monitor interval change such as temporal subtraction [64]: the main goal of screening is to assess progression of the disease and coal workers should be screened at least every five years. But no studies have investigated the analysis of radiographs of the same subject taken at different times.

Tully *et al.* [183], [184] were the first to focus on interstitial disease in general, using features from co-occurrence matrices to classify manually selected ROIs with linear discriminant analysis and sequential feature selection.

A large number of studies on the detection of interstitial disease has been performed by Katsuragawa and co-workers. The main elements of their approach are automatic selection of small ROIs within the costal and intercostal space [116], and computation of the standard deviation and the first moment of the Fourier spectrum [185], [186]. These two features are used to distinguish three classes of interstitial disease patterns (nodular, reticular, and honeycomb) from normal tissue. The same method has been applied to the standard radiographs from the International Labour Office for detection of pneumoconiosis in [187]. In other studies, geometric features from edges and ridges (line filters) and blobs (found by thresholding) were investigated [188], [189] and improved [190], and horizontal profiles of ROIs were used as input to a feed-forward neural network [191]. Features from several systems were combined and both thresholds and neural networks were used for classification. This hybrid scheme obtained a very high accuracy when tested on a database with 200 images from which 100 contained interstitial disease [191], [192]. The use of the scheme was also tested in an observer study [193].

A few studies on the detection of interstitial disease have been done by other groups. Kido *et al.* [194] computed geometric features, similar to the method later described in [189]. In another work, Kido *et al.* [195] classified interstitial abnormalities based on fractal analysis. Some papers deal with the effect of noise and blur in radiographic images on texture analysis methods such as features derived from co-occurrence matrices, the Fourier spectrum, morphological gradients, and fractal dimension [196], [197] and try to correct for these effects [198].

In [99], a scheme is presented in which texture feature vectors for different—overlapping—regions in the lung fields are compared with feature vectors from the same regions in other images. Thus, a large number of separate classifiers, one for each region, is used to locally analyze texture. The scheme is applied to a database from a tuberculosis mass screening program and to the database used by Katsuragawa and co-workers. The performance on this database is comparable with the results reported in [191]. The performance on the tuberculosis database was substantially lower. From the difference in performance of the same scheme on two databases we conclude once again that it is hard to compare results from different studies if different image sets are used. There are no established common databases to test the performance of computer algorithms for detecting interstitial disease. The results published so far seem to indicate that the detection of subtle interstitial abnormalities is still an open problem.

D. Other Tasks

There are other detection problems in chest radiography that do not fall into the previous sections. An example is the detection of pneumothorax. Sanada *et al.* [199] developed a method to detect pneumothoraces based on a Hough transform technique, evaluated on 50 images with 22 cases of pneumothorax. This is a difficult detection task that could be compared with detecting the tip of catheters [122]. Vogelsang *et al.* [76] present a method to detect catheters but do not present an evaluation.

V. DISCUSSION AND RECOMMENDATIONS FOR FUTURE RESEARCH

This survey gave an overview of three decades of research on computer analysis of chest radiographs in which over 150 papers on the subject have appeared. In this final section, we summarize the results by listing a number of solved and unsolved problems and possibilities for clinical applications by identifying some trends in the field, and with a list of recommendations for future research.

A. Solved and Unsolved Problems and Possibilities for Clinical Applications

- Several methods for segmentation of lung fields have obtained good results on test databases of reasonable size.
- Segmentation the heart and other large structures, such as the clavicles, has not received so much attention, but seems feasible as well.
- Automatic delineation of the posterior and anterior ribs is a harder problem for which no thoroughly evaluated methods have been proposed yet.
- Subtraction techniques (filtering out normal structures) have shown impressive results in distinct cases but are not yet applied in clinical practice.
- Most research has been focused at the detecting of lung nodules, and this has resulted in a number of schemes that have been evaluated on reasonably sized databases. For better performance in subtle cases, progress is needed.
- The results of texture analysis in chest radiograph are encouraging, but progress is needed to detect more subtle cases.

- Segmentation of branching blood vessels, and detection of objects such as clothing and catheters has not received much attention, although the results of such analyses could be used to eliminate false positives, to choose ROIs for texture analysis and to subtract normal structures. This remains an open problem.

The automatic detection of enlarged heart size and the total lung volume seem likely candidates for clinical application. Subtraction techniques seem also clinically applicable, especially for the detection of interval change.

However, lung nodule detection is the first area in which industry has developed computer-aided diagnosis products. This problem is difficult and far from solved, but the consequences of missed lung nodules spur several industries to develop products to assist radiologist in this task. Missed lung cancer is the second most common cause of malpractice suits among radiologist in the United States. In 90% of these cases the alleged error occurred on chest radiographs. About half of these cases resulted in indemnity payments averaging \$150 000 per case [200]. This year the first product for lung nodule detection, Rapidscreen by Deus Technologies, Rockville, MD, has received U.S. Food and Drug Administration marketing approval.

B. Trends

- The number of studies in which several methods are combined is increasing. Hybrid systems that use several types of features, sometimes including clinical information, are used to eliminate false positive nodules or to diagnose diseases.
- Radiology departments are starting to use central systems to store digital chest radiographs. This provides an opportunity to use "chest workstations" with intelligent algorithms that run in the background and may alert radiologists to possible abnormalities.
- So far, most research efforts in CAD have concentrated on the detection of possibly malignant structures, but recently the classification of abnormal signs is getting more attention. Experiments in which nodules are categorized as benign or malignant or in which signs of interstitial disease together with clinical findings are automatically diagnosed, have shown very promising results. Computer algorithms may become an important aid to the radiologist in differential diagnosis.
- We expect to see more applications of statistical knowledge-based techniques. Such methods are gaining popularity in many areas of medical image processing and the analysis of chest radiographs, where large numbers of images are available, seems ideally suited for their application.

C. Recommendations

- Focus on solving tasks that are encountered in clinical practice. Use data that contain abnormalities as they are encountered in practice, cf. the list given by MacMahon (see, [122] and Table III). The clinical experience with chest workstations should set the agenda for further research. CAD schemes should be developed and tested with radiographs as they are encountered in clinical practice. This is in contrast with previous research, in which the

focus always was on a single aspect or abnormality to be detected, and where algorithms were tested on a selected database with only those abnormalities and normal cases.

- Segmentation methods should preferably include quantitative results, offset against intraobserver and interobserver variability. The value of schemes that detect abnormalities can only be assessed when they are compared in observer studies with the performance of radiologists, or by comparing the performance of radiologists with and without the aid of a computer. In the latter case, the task for the radiologist should be similar to clinical practice.
- The use of dual energy subtraction images in computer analysis, that can be produced virtually for free by several commercially available chest units, could improve the sensitivity of CAD systems.
- Larger, publicly available databases should be collected and used for better validation. The work of Ho and Kruger [161] serves as an example. Research programs and funding agencies should become aware of this issue. The before mentioned chest workstations can be used for data collection over extended periods of time. Public databases will reduce the burden of individual research groups to collect data and allows fair comparisons of different techniques. So far, many papers report good or excellent results, but this can be deceiving because abnormalities in chest radiographs occur over such a wide range of subtleties. Increasing the size of databases is not only important for validation, but also for the development of better CAD schemes, that use data for training. Often the number of actually positive cases is limited compared with the number of features used for classification or feature selection or the number of parameters used in rule-based schemes.
- To increase the size of databases, it can be useful to explore the use of methods that use existing data to simulate new cases.
- Comparative studies are needed to assess the value of different algorithms. Researchers should be urged to make code implementations publicly available, to facilitate such studies.

ACKNOWLEDGMENT

This paper is a result of a four-year project on computer-aided diagnosis in chest radiography performed at the Image Sciences Institute, University Medical Center Utrecht, The Netherlands, which was supported by the Dutch Ministry of Economic Affairs through the IOP Image Processing program.

REFERENCES

- [1] W. Röntgen, Über eine neue art von strahlen, in *Sitzungsberichte der Physikalisch-Medicinisch Gesellschaft zu Würzburg*, pp. 132–141, 1895.
- [2] R. Daffner, *Clinical Radiology, the Essentials*, 2nd ed. Baltimore, MD: Williams & Wilkins, 1999.
- [3] R. Bunge and C. Herman, "Use of diagnostic imaging procedures: A nationwide hospital study," *Radiology*, vol. 163, no. 2, pp. 569–573, 1987.
- [4] L. Quekel, A. Kessels, R. Goei, and J. V. Engelshoven, "Miss rate of lung cancer on the chest radiograph in clinical practice," *Chest*, vol. 115, no. 3, pp. 720–724, 1999.

- [5] L. Garland, "On the scientific evaluation of diagnostic procedures," *Radiology*, vol. 52, no. 3, pp. 309–328, 1949.
- [6] J. Yerushalmy, "The statistical assessment of the variability in observer perception and description of Roentgenographic pulmonary shadows," *Radiologic Clinics N. Amer.*, vol. 7, no. 3, pp. 381–392, 1969.
- [7] H. Becker, W. Nettleton, P. Meyers, J. Sweeney, and C. Nice, Jr., "Digital computer determination of a medical diagnostic index directly from chest X-ray images," *IEEE Trans. Biomed. Eng.*, vol. BME-11, pp. 67–72, 1964.
- [8] P. Meyers, C. Nice, Jr., H. Becker, W. Nettleton, J. Sweeney, and G. Meckstroth, "Automated computer analysis of radiographic images," *Radiology*, vol. 83, pp. 1029–1034, 1964.
- [9] J. Duncan and N. Ayache, "Medical image analysis: Progress over two decades and the challenges ahead," *IEEE Trans. Pattern Anal. Machine Intell.*, vol. 22, pp. 85–106, Jan. 2000.
- [10] R. Connors, C. Harlow, and S. Dwyer III, "Radiographic image analysis: Past and present," in *Proc. 6th. Int. Conf. Pattern Recognition*, Munich, Germany, 1982, pp. 1152–1168.
- [11] R. Engle, Jr., "Attempts to use computers as diagnostic aids in medical decision making: A thirty-year experience," in *Perspectives Biol. Med.*, vol. 35, 1992, pp. 207–217.
- [12] M. Cocklin, A. Gourlay, P. Jackson, G. Kaye, I. Kerr, and P. Lams, "Digital processing of chest radiographs," *Image Vis. Comput.*, vol. 1, no. 2, pp. 67–78, 1983.
- [13] C. Schaefer-Prokop and M. Prokop, "Digital radiography of the chest: Comparison of the selenium detector with other imaging systems," *MedicaMundi*, vol. 41, no. 1, pp. 2–11, 1997.
- [14] J. Frija, E. D. Kerviler, A.-M. Zagdanski, C. Feger, P. Attal, and M. Laval-Jeantet, "Radiologie numérique du thorax," *Journal de Radiologie*, vol. 78, pp. 193–207, 1997.
- [15] L. Cook, G. Cox, M. Insana, M. McFadden, T. Hall, R. Gaborski, and F. Lure, "Comparison of a cathode-ray-tube and film for display of computed radiographic images," *Med Phys.*, vol. 25, no. 7, pp. 1132–1138, 1998.
- [16] J.-T. Ho and R. Kruger, "Comparison of dual and single exposure techniques in dual-energy chest radiography," *Med Phys.*, vol. 16, no. 2, pp. 202–208, 1989.
- [17] D. Ergun, W. Peppler, J. Dobbins III, F. Zink, D. Kruger, F. Kelcz, F. D. Bruijn, E. Bellers, Y. Wang, R. Althof, and M. Wind, "Dual-energy computer radiography," *Proc. SPIE*, vol. 2167, pp. 663–671, 1994.
- [18] D. Guant and G. Barnes, "X-ray tube potential, filtration, and detector considerations in dual-energy chest radiography," *Med Phys.*, vol. 21, no. 2, pp. 203–218, 1994.
- [19] Y. Asai, Y. Tanabe, Y. Ozaki, H. Kubota, M. Matsumoto, and H. Kanamori, "Optimum tube voltage for chest radiographs obtained by psychophysical analysis," *Med Phys.*, vol. 25, no. 11, pp. 2170–2175, 1998.
- [20] H. Kundel, C. Nodine, and E. Krupinsky, "Searching for lung nodules: Visual dwell indicates locations of false-positive and false-negative decisions," *Investigat. Radiol.*, vol. 24, pp. 424–427, 1989.
- [21] E. Krupinsky, C. Nodine, and H. Kundel, "A perceptually based method for enhancing pulmonary nodule recognition," *Investigat. Radiol.*, vol. 28, no. 4, pp. 289–294, 1993.
- [22] H. MacMahon, C. Vyborny, V. Sabeti, C. Metz, and K. Doi, "The effect of digital unsharp masking on the detectability of interstitial infiltrates and pneumothoraces," *Proc. SPIE*, vol. 555, pp. 246–252, 1985.
- [23] H. MacMahon, C. Vyborny, C. Metz, K. Doi, V. Sabeti, and S. Solomon, "Digital radiography of subtle pulmonary abnormalities: An ROC study of the effect of pixel size on observer performance," *Radiology*, vol. 158, pp. 21–26, 1986.
- [24] L.-N. Loo, K. Doi, and C. Metz, "Investigation of basic imaging properties in digital radiography: Effect of unsharp masking on the detectability of simple patterns," *Med Phys.*, vol. 12, no. 2, pp. 209–214, 1985.
- [25] R.-D. von Müller, H. Hirche, M. Voß, B. Buddenbrock, V. John, and P. Gocke, "ROC-Analyse zur Bildvernachlässigung digitaler Thoraxaufnahmen," *Fortschr. Röntgenstr.*, vol. 162, no. 2, pp. 163–169, 1995.
- [26] L. Cook, M. Insana, M. McFadden, T. Hall, and G. Cox, "Contrast-detail analysis of image degradation due to lossy compression," *Med Phys.*, vol. 22, no. 6, pp. 715–721, 1995.
- [27] G. Cox, L. Cook, M. Insana, M. McFadden, T. Hall, L. Harrison, D. Eckard, and N. Martin, "The effects of lossy compression on the detection of subtle pulmonary nodules," *Med Phys.*, vol. 23, no. 1, pp. 127–132, 1996.
- [28] E. Samei, "The Performance of Digital X-Ray Imaging Systems in Detection of Subtle Lung Nodules," Ph.D. dissertation, Univ. Michigan, Ann Arbor, May 1997.
- [29] G. Lodwick, T. Keats, and J. Dorst, "The coding of Roentgen images for computer analysis as applied to lung cancer," *Radiology*, vol. 81, no. 2, pp. 185–200, 1963.
- [30] A. Templeton, C. Jansen, J. Lehr, and R. Hufft, "Solitary pulmonary lesions," *Radiology*, vol. 89, no. 4, pp. 605–613, 1967.
- [31] N. Asada, K. Doi, H. MacMahon, S. Montner, M. Giger, C. Abe, and Y. Wu, "Potential usefulness of an artificial neural network for differential diagnosis of interstitial lung diseases: Pilot study," *Radiology*, vol. 177, no. 13, pp. 857–860, 1990.
- [32] K. Ashizawa, T. Ishida, H. MacMahon, C. Vyborny, S. Katsuragawa, and K. Doi, "Artificial neural networks in chest radiography," *Academic Radiol.*, vol. 6, pp. 2–9, 1999.
- [33] K. Ashizawa, H. MacMahon, T. Ishida, K. Nakamura, C. Vyborny, S. Katsuragawa, and K. Doi, "Effect of an artificial neural network on radiologists' performance in the differential diagnosis of interstitial lung disease using chest radiographs," *Amer. J. Roentgenol.*, vol. 172, pp. 1311–1315, 1999.
- [34] K. Abe, K. Doi, H. MacMahon, M. Giger, H. Jia, X. Chen, A. Kano, and T. Yanagisawa, "Computer-aided diagnosis in chest radiography, preliminary experience," *Investigat. Radiol.*, vol. 28, no. 11, pp. 987–993, 1993.
- [35] K. Doi, M. Giger, R. Nishikawa, K. Hoffmann, H. MacMahon, R. Schmidt, and K.-G. Chua, "Digital radiography: A useful clinical tool for computer-aided diagnosis by quantitative analysis of radiographic images," *Acta Radiologica*, vol. 34, pp. 426–439, 1993.
- [36] K. Doi, M. Giger, R. Nishikawa, K. Hoffmann, H. MacMahon, and R. Schmidt, "Potential usefulness of digital imaging in clinical diagnostic radiology: Computer-aided diagnosis," *J. Digital Imag.*, vol. 8, no. 1, pp. 2–7, 1995.
- [37] K. Doi, M. Giger, R. Nishikawa, K. Hoffmann, H. MacMahon, R. Schmidt, and C. Metz, "Recent progress in development of computer-aided diagnostic (CAD) schemes in radiology," *Med. Imag. Technol.*, vol. 13, no. 6, pp. 822–835, 1995.
- [38] K. Doi, "Perspectives on digital image analysis in medical imaging: Needs for a new science concerning technical understanding of the contents of medical images," *ICRU News*, vol. 1, pp. 10–14, June 1996.
- [39] K. Doi, H. MacMahon, S. Katsuragawa, R. Nishikawa, and Y. Jiang, "Computer-aided diagnosis in radiology: Potential and pitfalls," *Eur. J. Radiol.*, vol. 31, pp. 97–109, 1997.
- [40] K. Doi, *Computer-Aided Diagnosis in Medical Imaging*. Amsterdam, The Netherlands: Elsevier, 1999. Computer-aided diagnosis and its potential impact on diagnostic radiology.
- [41] *Handbook of Medical Imaging, 2: Medical Image Processing and Analysis*, M. Sonka and J. Fitzpatrick, Eds., SPIE Press, Bellingham, Wa, 2000, pp. 399–445. Medical image interpretation.
- [42] M. Giger and H. MacMahon, "Image processing and computer-aided diagnosis," *Radiologic Clinics N. Amer.*, vol. 34, no. 3, pp. 565–596, 1996.
- [43] H. Swett, M. Giger, and K. Doi, "Computer vision and decision support," in *Perception of Visual Information*, 2nd ed, W. Hendee and P. Wells, Eds. Berlin, Germany: Springer-Verlag, 1997, ch. 10, pp. 297–342.
- [44] K. Doi, H. MacMahon, M. Giger, and K. Hoffmann, Eds., *Computer-Aided Diagnosis in Medical Imaging*. Amsterdam, The Netherlands: Elsevier, 1999.
- [45] A. Reeves and W. Kostis, "Computer-aided diagnosis for lung cancer," *Radiologic Clinics N. Amer.*, vol. 38, no. 3, pp. 497–509, 2000.
- [46] C. Vyborny, M. Giger, and R. Nishikawa, "Computer-aided detection and diagnosis of breast cancer," *Radiologic Clinics N. Amer.*, vol. 38, no. 4, pp. 725–740, 2000.
- [47] *Handbook of Medical Imaging, 2: Medical Image Processing and Analysis*, M. Sonka and J. Fitzpatrick, Eds., SPIE Press, Bellingham, Wa, 2000, pp. 915–1004. Computer-aided diagnosis in mammography.
- [48] L. Squire and R. Novelline, *Fundamentals of Radiology*, 4th ed. Cambridge, MA: Harvard Univ. Press, 1988.
- [49] S. Lange, *Radiology of Chest Diseases*, Stuttgart: Georg Thieme Verlag, 1990.
- [50] R. Weissleder, M. Rieumont, and J. Wittenberg, *Primer of Diagnostic Imaging*, 2nd ed. St. Louis, MO: Mosby-Year Book, 1997.
- [51] W. Weiser, "Sorting out interstitial lung disease," *Appl. Radiol.*, pp. 58–65, Apr. 1993.
- [52] E. Hall, R. Kruger, S. Dwyer III, D. Hall, R. McLaren, and G. Lodwick, "A survey of preprocessing and feature extraction techniques for radiographic images," *IEEE Trans. Comput.*, vol. C-20, pp. 1032–1044, Sept. 1971.
- [53] R. Connors and C. Harlow, "Equal probability quantizing and texture analysis of radiographic images," *Comput. Graph. Image Processing*, vol. 8, pp. 447–463, 1978.
- [54] H. McAdams, G. Johnson, S. Suddarth, and C. Ravin, "Histograms-directed processing of digital chest images," *Investigat. Radiol.*, vol. 21, pp. 253–259, 1986.

- [55] S. Pizer, E. Amburn, J. Austin, R. Cromartie, A. Geselowitz, T. Greer, B. T. H. Romeny, J. Zimmerman, and K. Zuiderveld, "Adaptive histogram equalization and its variations," *Comput. Vis., Graph., Image Processing*, vol. 39, pp. 355–368, 1987.
- [56] R. Sherrier and G. Johnson, "Regionally adaptive histogram equalization of the chest," *IEEE Trans. Med. Imag.*, vol. 6, pp. 1–7, Feb. 1987.
- [57] M. I. Sezan, A. M. Tekalp, and R. Schaetzling, "Automatic anatomically selective image enhancement in digital chest radiography," *IEEE Trans. Med. Imag.*, vol. 8, pp. 154–161, Apr. 1989.
- [58] K. Rehm, G. Seeley, W. Dallas, R. Ovit, and J. Seeger, "Design and testing of artifact-suppressed adaptive histogram equalization: A contrast enhancement technique for display of digital chest radiographs," *J. Thoracic Imag.*, vol. 5, no. 1, pp. 85–91, 1990.
- [59] M. McNitt-Gray, R. Taira, S. Eldredge, and M. Razavi, "Brightness and contrast adjustments for different tissue densities in digital chest radiographs," *Proc. SPIE*, vol. 1445, pp. 468–478, 1991.
- [60] J.-P. Bolet, A. Cowen, J. Launder, A. Davies, and R. Bury, "Progress with an 'all-wavelet' approach to image enhancement and de-noising of direct digital thorax radiographic images," in *Proc. IPA'97*, 1997, pp. 244–248.
- [61] J. Suzuki, I. Furukawa, S. Ono, M. Kitamura, and Y. Ando, "Contrast mapping and evaluation for electronic X-ray images on CRT display monitor," *IEEE Trans. Med. Imag.*, vol. 16, pp. 772–784, June 1997.
- [62] J. Boone and J. Duryea, "Filter wheel equalization for chest radiography: A computer simulation," *Med Phys.*, vol. 22, no. 7, pp. 1029–1037, 1995.
- [63] R. Geluk, "Digital equalization radiography," *Proc. SPIE*, vol. 3659, pp. 471–477, 1999.
- [64] A. Kano, K. Doi, H. MacMahon, D. Hassell, and M. Giger, "Digital image subtraction of temporally sequential chest images for detection of interval change," *Med Phys.*, vol. 21, no. 3, pp. 453–461, 1994.
- [65] T. Ishida, K. Ashizawa, R. Engelmann, S. Katsuragawa, H. MacMahon, and K. Doi, "Application of temporal subtraction for detection of interval changes on chest radiographs: Improvement of subtraction images using automated initial image matching," *J. Digital Imag.*, vol. 12, no. 2, pp. 77–86, 1999.
- [66] T. Ishida, S. Katsuragawa, K. Nakamura, H. MacMahon, and K. Doi, "Iterative image warping technique for temporal subtraction of sequential chest radiographs to detect interval change," *Med Phys.*, vol. 26, pp. 1320–1329, 1999.
- [67] S. Katsuragawa, H. Tagashira, Q. Li, H. MacMahon, and K. Doi, "Comparison of the quality of temporal subtraction images obtained with manual and automated methods of digital chest radiography," *J. Digital Imag.*, vol. 12, no. 4, pp. 166–172, 1999.
- [68] M. Difazio, H. MacMahon, X.-W. Xu, P. Tsai, J. Shiraishi, S. Armato III, and K. Doi, "Digital chest radiography: Effect of temporal subtraction images on detection accuracy," *Radiology*, vol. 202, pp. 447–452, 1997.
- [69] H. Zhao, S. Lo, M. Freedman, and S. Mun, "On automatic temporal subtraction of chest radiographs and its enhancement for lung cancers," *Proc. SPIE*, vol. 4322, pp. 1867–1872, 2001.
- [70] Q. Li, S. Katsuragawa, T. Ishida, H. Yoshida, S. Tsukuda, H. MacMahon, and K. Doi, "Contralateral subtraction: A novel technique for detection of asymmetric abnormalities on digital chest radiographs," *Med Phys.*, vol. 27, no. 1, pp. 47–55, 2000.
- [71] Q. Li, S. Katsuragawa, and K. Doi, "Improved contralateral subtraction images by use of elastic matching technique," *Med Phys.*, vol. 27, no. 8, pp. 1934–1942, 2000.
- [72] H. Yoshida, "Multiresolution nonrigid image registration method and its application to removal of normal anatomical structures in chest radiographs," presented at the Int. Conf. Image Processing (ICIP'99, Kobe, Japan), Paper 27AP4.11, 1999.
- [73] H. Yoshida and K. Doi, "Computerized detection of pulmonary nodules in chest radiographs: Reduction of false positives based on symmetry between left and right lungs," *Proc. SPIE*, vol. 3979, pp. 97–102, 2000.
- [74] H. Yoshida, "Local contralateral subtraction based on simultaneous segmentation and registration method for computerized detection of pulmonary nodules," *Proc. SPIE*, vol. 4322, pp. 426–430, 2001.
- [75] F. Vogelsang, F. Weiler, J. Dahmen, M. Kilbinger, B. Wein, and R. Günther, "Detection and compensation of rib structures in chest radiographs for diagnose assistance," *Proc. SPIE*, vol. 3338, pp. 774–785, 1998.
- [76] F. Vogelsang, M. Kohnen, H. Schneider, F. Weiler, M. Kilbinger, B. Wein, and R. Günther, "Skeletal maturity determination from hand radiograph by model based analysis," *Proc. SPIE*, vol. 3979, pp. 294–305, 2000.
- [77] J. Toriwaki, Y. Suenaga, T. Negoro, and T. Fukumura, "Pattern recognition of chest X-ray images," *Comput. Graph. Image Processing*, vol. 2, pp. 252–271, 1973.
- [78] J. Toriwaki, J. Hasegawa, T. Fukumura, and Y. Takagi, "Computer analysis of chest photofluorograms and its applications to automated screening," *Automedica*, vol. 3, pp. 63–81, 1980.
- [79] C. Harlow and S. Eisenbeis, "The analysis of radiographic images," *IEEE Trans. Computers*, vol. 22, pp. 678–689, July 1973.
- [80] Y. Chien and K. Fu, "A decision function method for boundary detection," *Comput. Graph. Image Processing*, vol. 3, pp. 125–140, 1974.
- [81] —, "Recognition of X-ray picture patterns," *IEEE Trans. Syst., Man Cybern.*, vol. SMC-4, pp. 145–156, Feb. 1974.
- [82] A. Hasegawa, S. Lo, M. Freedman, and S. Mun, "Convolution neural network based detection of lung structures," *Proc. SPIE*, vol. 2167, pp. 654–662, 1994.
- [83] E. Pietka, "Lung segmentation in digital chest radiographs," *J. Digital Imag.*, vol. 2, pp. 79–84, 1994.
- [84] M. McNitt-Gray, J. Sayre, H. Huang, and M. Razavi, "A pattern classification approach to segmentation of chest radiographs," *Proc. SPIE*, vol. 1898, pp. 160–170, 1993.
- [85] M. McNitt-Gray, H. Huang, and J. Sayre, "Feature selection in the pattern classification problem of digital chest radiograph segmentation," *IEEE Trans. Med. Imag.*, vol. 14, pp. 537–547, June 1995.
- [86] J. Duryea and J. Boone, "A fully automatic algorithm for the segmentation of lung fields in digital chest radiographic images," *Med Phys.*, vol. 22, no. 2, pp. 183–191, 1995.
- [87] X. Xu and K. Doi, "Image feature analysis for computer-aided diagnosis: Accurate determination of ribcage boundary in chest radiographs," *Med Phys.*, vol. 22, no. 5, pp. 617–626, 1995.
- [88] —, "Image feature analysis for computer-aided diagnosis: Detection of right and left hemidiaphragm edges and delineation of lung field in chest radiographs," *Med Phys.*, vol. 23, no. 9, pp. 1613–1624, 1996.
- [89] S. Armato, M. Giger, and H. MacMahon, "Computerized delineation and analysis of costophrenic angles in digital chest radiographs," *Academic Radiol.*, vol. 5, pp. 329–335, 1998.
- [90] S. Armato, M. Giger, K. Ashizawa, and H. MacMahon, "Automated lung segmentation in digital lateral chest radiographs," *Med Phys.*, vol. 24, no. 8, pp. 1507–1520, 1998.
- [91] S. Armato, M. Giger, and H. MacMahon, "Automated lung segmentation in digitized postero-anterior chest radiographs," *Academic Radiol.*, vol. 4, pp. 245–255, 1998.
- [92] F. Carrascal, J. Carreira, M. Souto, P. Tahoces, L. Gomez, and J. Vidal, "Automatic calculation of total lung capacity from automatically traced lung boundaries in postero-anterior and lateral digital chest radiographs," *Med Phys.*, vol. 25, no. 7, pp. 1118–1131, 1998.
- [93] N. Vittitoe, R. Vargas-Voracek, and C. Floyd, Jr., "Identification of lung regions in chest radiographs using Markov Random Field modeling," *Med Phys.*, vol. 25, no. 6, pp. 976–985, 1998.
- [94] O. Tsujii, M. Freedman, and S. Mun, "Automated segmentation of anatomic regions in chest radiographs using an adaptive-sized hybrid neural network," *Med Phys.*, vol. 25, no. 6, pp. 998–1007, 1998.
- [95] L. Wilson, M. Brown, B. Doust, R. Gill, S. Brown, and D. Nair, "Computer-aided diagnosis using anatomical models," *Med. Imag. Technol.*, vol. 6, no. 14, pp. 652–663, 1996.
- [96] M. Brown, L. Wilson, B. Doust, R. Gill, and C. Sun, "Knowledge-based method for segmentation and analysis of lung boundaries in chest X-ray images," *Computerized Med. Imag. Graph.*, vol. 22, pp. 463–477, 1998.
- [97] N. Vittitoe, R. Vargas-Voracek, and C. Floyd, Jr., "Markov random field modeling in posteroanterior chest radiograph segmentation," *Med Phys.*, vol. 26, no. 8, pp. 1670–1677, 1999.
- [98] B. V. Ginneken and B. T. H. Romeny, "Automatic segmentation of lung fields in chest radiographs," *Med Phys.*, vol. 27, no. 10, pp. 2445–2455, 2000.
- [99] B. V. Ginneken, "Computer-Aided Diagnosis in Chest Radiography," Ph.D. dissertation, Utrecht Univ., Utrecht, The Netherlands, Mar. 2001.
- [100] T. Cootes, C. Taylor, D. Cooper, and J. Graham, "Active shape models—Their training and application," *Comput. Vis. Image Understanding*, vol. 61, no. 1, pp. 38–59, 1995.
- [101] R. Kruger, J. Townes, D. Hall, S. Dwyer III, and G. Lodwick, "Automated radiographic diagnosis via feature extraction and classification of cardiac size and shape descriptors," *IEEE Trans. Biomed. Eng.*, vol. BME-19, no. 3, pp. 174–186, 1972.
- [102] R. Sutton and E. Hall, "Texture measures for automatic classification of pulmonary disease," *IEEE Trans. Comput.*, vol. C-21, pp. 667–676, 1972.
- [103] J. Paul, M. Levine, R. Fraser, and C. Laszlo, "The measurement of total lung capacity based on a computer analysis of anterior and lateral radiographic chest images," *IEEE Trans. Biomed. Eng.*, vol. BME-21, pp. 444–451, June 1974.
- [104] E. Hall, W. Crawford, Jr., and F. Roberts, "Computer classification of pneumoconiosis from radiographs of coal workers," *IEEE Trans. Biomed. Eng.*, vol. BME-22, pp. 518–527, June 1975.
- [105] J. Jagoe and K. Paton, "Reading chest radiographs for pneumoconiosis by computer," *Br. J. Ind. Med.*, vol. 32, pp. 267–272, 1975.

- [106] H. Chotas and C. Ravin, "Chest radiography: Estimated lung volume and projected area obscured by the heart, mediastinum, and diaphragm," *Radiology*, vol. 193, no. 2, pp. 403–404, 1994.
- [107] H. Wechsler, *Automatic Detection of Rib Contours in Chest Radiographs*. New York: Birkhauser Verlag, 1977.
- [108] H. Wechsler and J. Sklansky, "Finding the rib cage in chest radiographs," *Pattern Recogn.*, vol. 9, pp. 21–30, 1977.
- [109] H. Wechsler, "Image processing algorithms applied to rib boundary detection in chest radiographs," *Comput. Graph. Image Processing*, vol. 7, pp. 375–390, 1978.
- [110] D. Ballard, "Model-directed detection of ribs in chest radiographs," in *Proc. 4th Int. Joint Conf. Pattern Recognition*, 1978, pp. 907–910.
- [111] P. D. Souza, "Automatic rib detection in chest radiographs," *Comput. Vis. Graph. Image Processing*, vol. 23, pp. 129–161, 1983.
- [112] G. Powell, K. Doi, and S. Katsuragawa, "Localization of inter-rib spaces for lung texture analysis and computer-aided diagnosis in digital chest images," *Med Phys.*, vol. 15, no. 4, pp. 581–587, 1988.
- [113] S. Sanada, K. Doi, and H. MacMahon, "Image feature analysis and computer-aided diagnosis in digital radiography: Automated delineation of posterior ribs in chest images," *Med Phys.*, vol. 18, no. 5, pp. 964–971, 1991.
- [114] Z. Yue, A. Goshtasby, and L. Ackerman, "Automatic detection of rib borders in chest radiographs," *IEEE Trans. Med. Imag.*, vol. 14, pp. 525–536, June 1995.
- [115] B. V. Ginneken and B. T. H. Romeny, "Automatic delineation of ribs in chest radiographs," *Proc. SPIE*, vol. 3979, pp. 825–836, 2000.
- [116] X. Chen, K. Doi, S. Katsuragawa, and H. MacMahon, "Automated selection of regions of interest for quantitative analysis of lung textures in digital chest radiographs," *Med Phys.*, vol. 20, no. 4, pp. 975–982, 1993.
- [117] D. Hall, G. Lodwick, R. Kruger, and S. Dwyer III, "Computer diagnosis of heart disease," *Radiological Clinics N. Amer.*, vol. 9, no. 3, pp. 533–541, 1971.
- [118] N. Sezaki and K. Ukena, "Automatic computation of the cardiothoracic ratio with application to mass screening," *IEEE Trans. Biomed. Eng.*, vol. BME-20, pp. 248–253, Apr. 1973.
- [119] N. Nakamori, K. Doi, V. Sabeti, and H. MacMahon, "Image feature analysis and computer-aided diagnosis in digital radiography: Automated analysis of sizes of heart and lung in chest images," *Med Phys.*, vol. 17, no. 3, pp. 342–350, 1990.
- [120] N. Nakamori, K. Doi, H. MacMahon, Y. Sasaki, and S. Montner, "Effect of heart-size parameters computed from digital chest radiographs on detection of cardiomegaly: Potential usefulness for computer-aided diagnosis," *Investigat. Radiol.*, vol. 26, no. 6, pp. 546–550, 1991.
- [121] S. Armato, M. Giger, and H. MacMahon, "Computerized detection of abnormal asymmetry in digital chest radiographs," *Med Phys.*, vol. 21, no. 11, pp. 1761–1768, 1994.
- [122] H. MacMahon, K. Liu, S. Montner, and K. Doi, "The nature and subtlety of abnormal findings in chest radiographs," *Med Phys.*, vol. 18, no. 2, pp. 206–210, 1991.
- [123] F. Roellinger, A. Kahveci, J. Chang, C. Harlow, S. Dwyer III, and G. Lodwick, "Computer analysis of radiographic images," *Comput. Graph. Image Processing*, vol. 2, pp. 232–251, 1973.
- [124] J. Sklansky and D. Ballard, "Tumor detection in radiographs," *Comput. Biomed. Res.*, vol. 6, no. 4, pp. 299–321, 1973.
- [125] D. Ballard, *Hierarchical Recognition of Tumors in Chest Radiographs*. New York: Birkhauser Verlag, 1976.
- [126] D. Ballard and J. Sklansky, "A ladder-structured decision tree for recognizing tumors in chest radiographs," *IEEE Trans. Comput.*, vol. C-20, pp. 503–513, 1976.
- [127] J. Sklansky and D. Petković, "Two-resolution detection of lung tumors in chest radiographs," in *Multiresolution Image Processing and Analysis*, A. Rosenfeld, Ed. Berlin, Germany: Springer-Verlag, 1984, pp. 365–378.
- [128] W. Lampeter and J. Wandtke, "Computerized search of chest radiographs for nodules," *Investigat. Radiol.*, vol. 21, pp. 384–390, 1986.
- [129] M. Giger, K. Doi, and H. MacMahon, "Image feature analysis and computer-aided diagnosis in digital radiography: Automated detection of nodules in peripheral lung fields," *Med Phys.*, vol. 15, no. 2, pp. 158–166, 1988.
- [130] H. Yoshimura, M. Giger, K. Doi, H. MacMahon, and S. Montner, "Computerized scheme for the detection of pulmonary nodules: A nonlinear filtering technique," *Investigat. Radiol.*, vol. 27, pp. 124–127, 1992.
- [131] H. Suzuki, N. Inaoka, H. Takabatake, M. Mori, H. Natori, and A. Suzuki, "An experimental system for detecting lung nodules by chest X-ray image processing," *Proc. SPIE*, vol. 1450, pp. 99–107, 1991.
- [132] H. Suzuki and N. Inaoka, "Development of a computer-aided detection system for lung cancer diagnosis," *Proc. SPIE*, vol. 1652, pp. 567–571, 1992.
- [133] S.-C. Lo, M. Freedman, J.-S. Lin, and S. Mun, "Automatic lung nodule detection using profile matching and back-propagation neural network techniques," *J. Digital Imag.*, vol. 6, no. 1, pp. 48–54, 1993.
- [134] H. Yoshida, X. Xu, K. Doi, and M. Giger, "Computer-aided diagnosis (CAD) scheme for detecting pulmonary nodules using wavelet transforms," *Proc. SPIE*, vol. 2434, pp. 621–626, 1995.
- [135] F. Mao, W. Qian, J. Gaviria, and L. Clarke, "Fragmentary window filtering for multiscale lung nodule detection," *Academic Radiol.*, vol. 5, no. 4, pp. 306–311, 1998.
- [136] J. Wei, Y. Hagihara, and H. Kobatake, "Detection of cancerous tumors on chest X-ray images—Candidate detection filter and its evaluation," presented at the Int. Conf. Image Processing (ICIP'99), Paper 27AP4.2, 1999.
- [137] P. Sankar and J. Sklansky, "A Gestalt guided heuristic boundary follower for X-ray images of lung nodules," *IEEE Trans. Pattern Anal. Machine Intell.*, vol. PAMI-4, pp. 326–331, Mar. 1982.
- [138] W. Lampeter, "ANDS-V1 computer detection of lung nodules," *Proc. SPIE*, vol. 555, pp. 253–261, 1985.
- [139] M. Giger, K. Doi, H. MacMahon, C. Metz, and F.-F. Yin, "Pulmonary nodules: Computer-aided detection in digital chest images," *Radiographics*, vol. 10, pp. 41–51, 1990.
- [140] G. Cox, F. Hoare, and G. D. Jager, "Experiments in lung cancer nodule detection using texture analysis and neural network classifiers," presented at the 3rd S. African Workshop Pattern Recognition, 1992.
- [141] R. Haralick, K. Shanmugam, and I. Dinstein, "Textural features for image classification," *IEEE Trans. Syst., Man Cybern.*, vol. SMC-3, pp. 610–621, 1973.
- [142] K. Laws, "Texture image segmentation," Ph.D. dissertation, Univ. Southern Calif., Los Angeles, 1980.
- [143] T. Matsumoto, H. Yoshimura, K. Doi, M. Giger, A. Kano, H. MacMahon, M. Abe, and S. Montner, "Image feature analysis of false-positive diagnoses produced by automated detection of lung nodules," *Investigat. Radiol.*, vol. 27, no. 8, pp. 587–579, 1992.
- [144] Y. P. Chiou, J.-S. Lin, Y. F. Lure, P. Ligomenides, M. Freedman, and S. Fritz, "Shape feature analysis using artificial neural networks for improvements of hybrid lung nodule detection (hlnsd) system," *Proc. SPIE*, vol. 1898, pp. 609–617, 1993.
- [145] Y. Wu, K. Doi, M. Giger, C. Metz, and W. Zhang, "Reduction of false positives in computerized detection of lung nodules in chest radiographs using artificial neural networks, discriminant analysis, and a rule-based scheme," *J. Digital Imag.*, vol. 7, no. 4, pp. 196–207, 1994.
- [146] S.-C. Lo, S.-L. Lou, J.-S. Lin, M. Freedman, M. Chien, and S. Mun, "Artificial convolution neural network techniques and applications for lung nodule detection," *IEEE Trans. Med. Imag.*, vol. 14, pp. 711–718, Aug. 1995.
- [147] J.-S. Lin, A. Hasegawa, M. Freedman, and S. Mun, "Differentiation between nodules and end-on vessels using a convolution neural network architecture," *J. Digital Imag.*, vol. 8, no. 3, pp. 132–141, 1995.
- [148] J.-S. Lin, S. Lo, A. Hasegawa, M. Freedman, and S. Mun, "Reduction of false positives in lung nodule detection using a two-level neural classification," *IEEE Trans. Med. Imag.*, vol. 15, pp. 206–217, Feb. 1996.
- [149] C. Floyd, E. Patz, J. Lo, N. Vittitoe, and L. Stambaugh, "Diffuse nodular lung disease on chest radiographs: A pilot study of characterization by fractal dimension," *Amer. J. Roentgenol.*, vol. 167, pp. 1185–1187, 1996.
- [150] N. Vittitoe, J. Baker, and C. Floyd, Jr., "Fractal texture analysis in computer-aided diagnosis of solitary pulmonary nodules," *Academic Radiol.*, vol. 4, pp. 96–101, 1997.
- [151] X. Xu, H. MacMahon, M. Giger, and K. Doi, "Adaptive feature analysis of false positives for computerized detection of lung nodules in digital chest radiographs," *Proc. SPIE*, vol. 3034, pp. 428–436, 1997.
- [152] M. Carreira, D. Cabello, M. Penedo, and A. Mosquera, "Computer-aided diagnoses: Automatic detection of lung nodules," *Med Phys.*, vol. 25, no. 10, pp. 1998–2006, 1998.
- [153] M. Penedo, M. Carreira, A. Mosquera, and D. Cabello, "Computer-aided diagnosis: A neural-network-based approach to lung nodule detection," *IEEE Trans. Med. Imag.*, vol. 17, pp. 872–880, June 1998.
- [154] X. Xu, S. Katsuragawa, K. Ashizawa, H. MacMahon, and K. Doi, "Analysis of image features of histograms of edge gradient for false positive reduction in lung nodule detection in chest radiographs," *Proc. SPIE*, vol. 3338, pp. 318–326, 1998.
- [155] M. Casaldi, G. Russo, G. Scarano, and P. Talone, "Automatic detection of lung nodules: Application to radiograms lossy coding," *Proc. SPIE*, vol. 3338, pp. 1365–1376, 1998.
- [156] K. Nakamura, H. Yoshida, R. Engelmann, H. MacMahon, S. Katsuragawa, T. Ishida, K. Ahizawa, and K. Doi, "Computerized analysis of the likelihood of malignancy in solitary pulmonary nodules with use of artificial neural networks," *Radiology*, vol. 214, pp. 823–830, 2000.

- [157] D. Catarious, Jr., A. Baydush, and C. Floyd, Jr., "Initial development of a computer-aided diagnosis tool for solitary pulmonary nodules," *Proc. SPIE*, vol. 4322, pp. 710–717, 2001.
- [158] Q. Li, S. Katsuragawa, R. Engelmann, S. Armoto, H. MacMahon, and K. Doi, "Development of a multiple-templates matching technique for removal of false positives in computer-aided diagnostic scheme," *Proc. SPIE*, vol. 4322, pp. 1763–1770, 2001.
- [159] J. Muhm, W. Miller, R. Fontana, D. Sanderson, and M. Uhlenhopp, "Lung cancer detected during a screening program using four-month chest radiographs," *Radiology*, vol. 148, pp. 609–615, 1983.
- [160] A. Jain, *Fundamentals of Digital Image Processing*. Englewood Cliffs, NJ: Prentice Hall, 1989.
- [161] J. Shiraishi, S. Katsuragawa, J. Ikezoe, T. Matsumoto, T. Kobayashi, K. Komatsu, M. Matsui, H. Fujita, Y. Kodera, and K. Doi, "Development of a digital image database for chest radiographs with and without a lung nodule: Receiver operating characteristic analysis of radiologists' detection of pulmonary nodules," *Amer. J. Roentgenol.*, vol. 174, pp. 71–74, 2000.
- [162] R. Nishikawa, M. Giger, K. Doi, C. Metz, F.-F. Yin, V. C. J. <Author: Please supply last name?>, and R. Schmidt, "Effect of case selection of the performance of computer-aided detection schemes," *Med Phys.*, vol. 21, no. 2, pp. 265–269, 1994.
- [163] C. Metz, "ROC methodology in radiologic imaging," *Investigat. Radiol.*, vol. 21, no. 9, pp. 720–733, 1986.
- [164] —, "Some practical issues of experimental design and data analysis in radiological ROC studies," *Investigat. Radiol.*, vol. 24, no. 3, pp. 234–245, 1989.
- [165] J. Gurney and S. Swensen, "Solitary pulmonary nodules: Determining the likelihood of malignancy with neural network analysis," *Radiology*, vol. 196, pp. 823–829, 1995.
- [166] T. Matsumoto, H. Yoshimura, M. Giger, K. Doi, A. Kano, H. MacMahon, M. Abe, and S. Montner, "Potential usefulness of computerized nodule detection in screening programs for lung cancer," *Investigat. Radiol.*, vol. 27, pp. 471–475, 1992.
- [167] T. Kobayashi, X.-W. Xu, H. MacMahon, C. Metz, and K. Doi, "Effect of a computer-aided diagnosis scheme on radiologists' performance in detection of lung nodules on radiographs," *Radiology*, vol. 199, pp. 843–848, 1996.
- [168] H. MacMahon, R. Engelmann, F. Behlen, K. Hoffmann, T. Ishida, C. Roe, C. Metz, and K. Doi, "Computer-aided diagnosis of pulmonary nodules: Results of a large-scale observer test," *Radiology*, vol. 213, pp. 723–726, 1999.
- [169] H. Kundel, G. Revesz, and L. Toto, "Contrast gradient and the detection of lung nodules," *Investigat. Radiol.*, vol. 14, pp. 18–22, 1979.
- [170] L. Quekel, "Detectability of early lung cancer on the chest radiograph: A study on miss rate and observer performance in clinical practice," Ph.D. dissertation, Maastricht Univ., <Location?>, 2001.
- [171] R. Sherrier, G. Johnson, S. Suddarth, C. Chiles, C. Hulka, and C. Ravin, "Digital synthesis of lung nodules," *Investigat. Radiol.*, vol. 20, no. 9, pp. 933–937, 1985.
- [172] R. Haralick, "Statistical and structural approaches to texture," *Proc. IEEE*, vol. 67, pp. 786–804, May 1979.
- [173] G. Revesz and H. Kundel, "Feasibility of classifying disseminated pulmonary diseases based on their Fourier spectra," *Investigat. Radiol.*, vol. 8, pp. 345–349, 1973.
- [174] H. Stark and D. Lee, "An optical-digital approach to the pattern recognition of coal workers' pneumoconiosis," *IEEE Trans. Syst., Man Cybern.*, vol. SMC-6, pp. 788–793, Nov. 1976.
- [175] R. Kruger, W. Thompson, and A. Turner, "Computer diagnosis of pneumoconiosis," *IEEE Trans. Syst., Man, Cybern.*, vol. SMC-4, p. 40, 1974.
- [176] A. Turner, R. Kruger, and W. Thompson, "Automated computer screening of chest radiographs for pneumoconiosis," *Investigat. Radiol.*, vol. 11, no. 4, pp. 258–266, 1976.
- [177] E. Hall, R. Kruger, and A. Turner, "An optical-digital system for automatic processing of chest X-rays," *Opt. Eng.*, vol. 13, no. 3, pp. 250–257, 1974.
- [178] R. Ledley, H. Huang, and L. Rotola, "A texture analysis method in classification of coal workers' pneumoconiosis," *Comput. Biol. Med.*, vol. 5, pp. 53–67, 1975.
- [179] J. Jagoe and K. Paton, "Measurement of pneumoconiosis by computer," *IEEE Trans. Comput.*, vol. C-25, pp. 95–97, 1976.
- [180] J. Jagoe, "Gradient pattern coding—An application to the measurement of pneumoconiosis in chest X-rays," *Comput. Biomed. Res.*, vol. 12, pp. 1–15, 1979.
- [181] C. Li and A. Savol, "Computer analysis of small opacities in coal miners' chest X-rays," *Proc. IEEE Int. Conf. Cybernetics and Society*, pp. 422–428, 1977.
- [182] P. Soliz, M. Pattichis, J. Ramachandran, and D. James, "Computer-assisted diagnosis of chest radiographs for pneumoconioses," in *SPIE*, vol. 4322, 2001, pp. 667–657.
- [183] R. Tully, R. Connors, C. Harlow, G. Larsen, S. Dwyer, III, and G. Lodwick, "Interactive analysis of pulmonary infiltration," in *Proc. 3rd Int. Joint Conf. Pattern Recognition*, 1976, pp. 238–242.
- [184] R. Tully, R. Connors, C. Harlow, and G. Lodwick, "Toward computer analysis of pulmonary infiltration," *Investigat. Radiol.*, vol. 13, pp. 298–305, 1978.
- [185] S. Katsuragawa, K. Doi, and H. MacMahon, "Image feature analysis and computer-aided diagnosis in digital radiography: Detection and characterization of interstitial lung disease in digital chest radiographs," *Med Phys.*, vol. 15, no. 3, pp. 311–319, 1988.
- [186] —, "Image feature analysis and computer-aided diagnosis in digital radiography: Classification of normal and abnormal lungs with interstitial disease in chest images," *Med Phys.*, vol. 16, no. 1, pp. 38–44, 1989.
- [187] S. Katsuragawa, K. Doi, H. MacMahon, N. Nakamori, Y. Sasaki, and J. Fennessy, "Quantitative computer-aided analysis of lung texture in chest radiographs," *Radiographics*, vol. 10, no. 2, pp. 257–269, 1990.
- [188] L. Monnier-Cholley, H. MacMahon, S. Katsuragawa, J. Morishita, T. Ishida, and K. Doi, "Computerized analysis of interstitial infiltrates on chest radiographs: A new scheme based on geometric pattern features and Fourier analysis," *Academic Radiol.*, vol. 2, pp. 455–462, 1995.
- [189] S. Katsuragawa, K. Doi, H. MacMahon, L. Monnier-Cholley, J. Morishita, and T. Ishida, "Quantitative analysis of geometric pattern features of interstitial infiltrates in digital chest radiographs: Preliminary results," *J. Digital Imag.*, vol. 9, no. 3, pp. 137–144, 1996.
- [190] T. Ishida, S. Katsuragawa, T. Kobayashi, H. MacMahon, and K. Doi, "Computerized analysis of interstitial disease in chest radiographs: Improvement of geometric-pattern feature analysis," *Med Phys.*, vol. 24, no. 6, pp. 915–924, 1997.
- [191] T. Ishida, S. Katsuragawa, K. Ashizawa, H. MacMahon, and K. Doi, "Application of artificial neural networks for quantitative analysis of image data in chest radiographs for detection of interstitial lung disease," *J. Digital Imag.*, vol. 11, no. 4, pp. 182–192, 1998.
- [192] S. Katsuragawa, K. Doi, H. MacMahon, L. Monnier-Cholley, T. Ishida, and T. Kobayashi, "Classification of normal and abnormal lungs with interstitial diseases by rule-based method and artificial neural networks," *J. Digital Imag.*, vol. 10, no. 3, pp. 108–114, 1997.
- [193] L. Monnier-Cholley, H. MacMahon, S. Katsuragawa, J. Morishita, T. Ishida, and K. Doi, "Computer-aided diagnosis for detection of interstitial opacities on chest radiographs," *Amer. J. Roentgenol.*, vol. 171, pp. 1651–1656, 1998.
- [194] S. Kido, J. Ikezoe, H. Naito, M. Masuie, S. Tamura, and T. Kozuka, "An image analyzing system for interstitial lung abnormalities in chest radiography," *Radiographics*, vol. 15, pp. 1457–1464, 1995.
- [195] S. Kido, J. Ikezoe, H. Naito, S. Tamura, and S. Machi, "Fractal analysis of interstitial lung abnormalities in chest radiography," *Radiographics*, vol. 15, pp. 1457–1464, 1995.
- [196] J. Veenland, J. Grashuis, and E. Gelsema, "Texture analysis in radiographs: The influence of modulation transfer function and noise on the discriminative ability of texture features," *Med Phys.*, vol. 25, no. 6, pp. 922–936, 1998.
- [197] J. Veenland, J. Grashuis, F. V. D. Meer, A. Beckers, and E. Gelsema, "Estimation of fractal dimension in radiographs," *Med Phys.*, vol. 23, no. 4, pp. 585–594, 1996.
- [198] J. Morishita, K. Doi, S. Katsuragawa, L. Monnier-Cholley, and H. MacMahon, "Computer-aided diagnosis for interstitial infiltrates in chest radiographs: Optical-density dependence of texture measures," *Med Phys.*, vol. 22, no. 9, pp. 1515–1522, 1995.
- [199] S. Sanada, K. Doi, and H. MacMahon, "Image feature analysis and computer-aided diagnosis in digital radiography: Automated detection of pneumothorax in chest images," *Med Phys.*, vol. 19, no. 5, pp. 1153–1160, 1992.
- [200] C. White and C. Meyer, "Missed lung cancer: Medicolegal implications," *Appl. Radiol.*, vol. 27, no. 8, 1998.