

In-vitro and in-silico porous phantoms for investigating the relationship between microvascular architecture and ultrasound-contrast-agent kinetics

Citation for published version (APA):

Chen, P., Turco, S., van Sloun, R. J. G., Pollet, A., den Toonder, J. M. J., Wijkstra, H., & Mischi, M. (2020). In-vitro and in-silico porous phantoms for investigating the relationship between microvascular architecture and ultrasound-contrast-agent kinetics. In *Proceedings of the IEEE International Ultrasonics Symposium (IUS)* Institute of Electrical and Electronics Engineers.

Document status and date:

Published: 01/01/2020

Document Version:

Accepted manuscript including changes made at the peer-review stage

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.tue.nl/taverne

Take down policy

If you believe that this document breaches copyright please contact us at:

openaccess@tue.nl

providing details and we will investigate your claim.

***In-vitro* and *in-silico* porous phantoms for investigating the relationship between microvascular architecture and ultrasound-contrast-agent kinetics**

Peiran Chen¹, Simona Turco¹, Ruud van Sloun¹, Andreas Pollet², Jaap den Toonder², Hessel Wijkstra^{1,3}, Massimo Mischi¹

¹Department of Electrical Engineering, Eindhoven University of Technology, Netherlands,

²Department of Mechanical Engineering, Eindhoven University of Technology, Netherlands,

³Department of Urology, Amsterdam University Medical Centers, Netherlands

Background, Motivation and Objective

As a recognized hallmark of cancer, tumor-driven angiogenic microvasculature is characterized by increased microvascular density (MVD), smaller vessel diameter (d_v), and higher vessel tortuosity, leading to complex blood flow patterns [1, 2]. Analysis of the dispersion kinetics of an ultrasound contrast agent (UCA) in the tumor vasculature has shown promise for prostate cancer (PCa) diagnostics [1, 2], but the link between UCA kinetics and underlying microvascular architecture is still under investigation. In previous work, modeling the microvasculature as a porous medium, we developed *in-vitro* phantoms and imaged their perfusion by contrast enhanced ultrasound. The experimental results were consistent with our *in-vivo* findings [3]. In this work, we developed a simulation framework to further advance our understanding of the UCA kinetics through porous media that mimic microvascular networks.

Statement of Contribution/Methods

The *in-silico* model of a porous phantom was realized using a 3D sphere-packing algorithm. Thousands of mono-sized spheres were uniformly distributed inside a cylindrical space. Spheres were then separated iteratively until no overlapping was present. For each 100 iterations, random motion was applied to all the spheres mimicking a shaking process. In agreement with our previous *in-vitro* measurements, we simulated porous phantoms with sphere diameters of 3.1, 2.5, and 1.6 mm. The distribution of the radial porosity within the cylinder was obtained by calculating the void fraction in each concentric layer of 0.04-mm thickness. Moreover, Delaunay triangulation was applied to the simulated phantoms, enabling us to measure the pore size represented by the size of the inscribed spheres of each connected tetrahedron. Furthermore, the MVD was reflected by the number of pores. Dynamic simulations of bubble transport flowing through the porous phantoms are being tested by COMSOL.

Results/Discussion

The simulated porous phantoms are shown in Fig. 1. The pore size and density distribution demonstrate that the pore size decreases while the density increases as the bead size decreases, representing smaller d_v and higher MVD. This result confirms our hypothesis that porous phantoms with smaller beads can mimic angiogenic microvasculature.

