

Contrast ultrasound dispersion imaging in prostate cancer diagnostics

Citation for published version (APA):

Mischi, M., Kuenen, M. P. J., Laguna, M. P., Rosette, de la, J. J. M. C. H., & Wijkstra, H. (2013). Contrast ultrasound dispersion imaging in prostate cancer diagnostics. In *Proceedings of the 28th Annual Meeting of the Engineering & Urology Society, 4 May 2013, San Diego, USA* (pp. 89-89)

Document status and date:

Published: 01/01/2013

Document Version:

Publisher's PDF, also known as Version of Record (includes final page, issue and volume numbers)

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.tue.nl/taverne

Take down policy

If you believe that this document breaches copyright please contact us at:

openaccess@tue.nl

providing details and we will investigate your claim.

CONTRAST ULTRASOUND DISPERSION IMAGING IN PROSTATE CANCER DIAGNOSTICS

M. Mischi¹, M.P.J. Kuenen^{1,2}, M.P. Laguna Pes², J.J.M.C.H. de la Rosette², H. Wijkstra^{1,2}

¹*Eindhoven University of Technology, Electrical Engineering Department, the Netherlands*

²*Academic Medical Center, University of Amsterdam, Urology Department, the Netherlands*

Introduction: In the United States, prostate cancer (PCa) accounts for 28% and 10% of all cancer diagnoses and deaths in males, respectively. Although efficient focal therapies are available, their use is hampered by a lack of reliable imaging that enables accurate PCa localization. Contrast-ultrasound dispersion imaging (CUDI) has been proposed as a new alternative method for PCa localization based on dynamic contrast-enhanced ultrasound (DCE-US) data. Different from other DCE-US methods for cancer localization, invariably based on the assessment of blood perfusion, the intravascular dispersion of ultrasound contrast agents is directly influenced by the microvascular changes produced by those angiogenic processes supporting cancer growth. Characterization of the microvascular architecture opens therefore new possibilities for noninvasive assessment of PCa aggressiveness (cancer grading).

Methods: CUDI is performed after an intravenous injection of a 2.4-mL SonoVue[®] (Bracco, Milan) bolus. The bolus passage through the prostate is imaged by an iU22 ultrasound scanner (Philips Healthcare, Bothell) equipped with a transrectal probe (C8-4v or C10-3v). Contrast sensitivity is increased by employment of contrast-specific imaging. Time concentration curves (TCCs) are obtained at each video pixel after data linearization. Local dispersion is analyzed by assessment of the shape similarity between neighbor TCCs. As shown in Figure 1, a parametric dispersion map is generated that highlights angiogenic areas (in red color) and, therefore, aggressive cancer areas. A preliminary validation was performed at the Academic Medical Center University of Amsterdam by comparing the obtained dispersion images with the histology results after radical prostatectomy (Figure 1) in 12 datasets recorded from 8 patients.

Results: In all patients, the dispersion images showed a good agreement on a pixel level with the histology. The resulting average receiver operating characteristic (ROC) curve area was 0.89, with sensitivity and specificity equal to 77.3% and 86.0%, respectively. These results were superior to those obtained by all the tested perfusion parameters which were previously proposed in the literature.




Figure: Dispersion map (a) with corresponding histology result (b).

Conclusion: Our preliminary results evidence a promising value of CUDI for PCa localization and motivate towards an extensive validation. Once validated, CUDI could support targeting of biopsy and focal-therapy, as well as therapeutic decision making based on noninvasive cancer grading.